

LB116023


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

44962



A= 0.0000e+000

B= 2.2271e-004 slope

C= -1.7500e-002 intercept

Rho= 0.9999805

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%ob
0.00	0.000	0.022	0.022	176	0.000	176					0
0.05	0.050					573					-4
0.20	0.200	0.193	-0.007	947	0.0 %	947					-2
2.50	2.500	2.460	-0.040	11123	0.0 %	11123					1
5.00	5.000	5.032	0.032	22673	0.0 %	22673					0
7.50	7.500	7.497	-0.003	33740	0.0 %	33740					0
10.0	10.000	9.996	-0.004	44962	0.0 %	44962					0

LB116023
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	176	-		0 SFAM01.1	PPB	8/25/2021 13:45	Std
	0.2 S01	947	-		0.2 SFAM01.1	PPB	8/25/2021 13:47	Std
	2.5 S02	11123	-		2.5 SFAM01.1	PPB	8/25/2021 13:50	Std
	5 S03	22673	-		5 SFAM01.1	PPB	8/25/2021 13:52	Std
	7.5 S04	33740	-		7.5 SFAM01.1	PPB	8/25/2021 13:54	Std
	10 S05	44962	-		10 SFAM01.1	PPB	8/25/2021 13:59	Std
ICV061	ICV061	17680	3.9201 -		SFAM01.1	PPB	8/25/2021 14:02	SMPL
ICB061	ICB061	115	0.0081 -		SFAM01.1	PPB	8/25/2021 14:04	SMPL
CCV042	CCV042	22528	4.9998 -		SFAM01.1	PPB	8/25/2021 14:06	SMPL
CCB042	CCB042	126	0.0106 -		SFAM01.1	PPB	8/25/2021 14:08	SMPL
PB138704BL	PBS704	6	-0.0162 -		SFAM01.1	PPB	8/25/2021 14:13	SMPL
M3460-01	MCOAR3	2252	0.4841 -		SFAM01.1	PPB	8/25/2021 14:15	SMPL
M3460-02	MCOAR4	1096	0.2266 -		SFAM01.1	PPB	8/25/2021 14:18	SMPL
M3460-03	MCOAR7	1690	0.3589 -		SFAM01.1	PPB	8/25/2021 14:20	SMPL
M3460-04	MCOAS1	1466	0.309 -		SFAM01.1	PPB	8/25/2021 14:22	SMPL
M3460-05	MCOAS2	4089	0.8932 -		SFAM01.1	PPB	8/25/2021 14:24	SMPL
M3460-06	MCOAT8	682	0.1344 -		SFAM01.1	PPB	8/25/2021 14:27	SMPL
M3460-07	MCOAW3	1333	0.2794 -		SFAM01.1	PPB	8/25/2021 14:29	SMPL
M3460-08	MCOAW7	981	0.201 -		SFAM01.1	PPB	8/25/2021 14:31	SMPL
M3460-09	MCOAW8	3113	0.6758 -		SFAM01.1	PPB	8/25/2021 14:34	SMPL
M3460-10	COB56	1945	0.4157 -		SFAM01.1	PPB	8/25/2021 14:36	SMPL
M3460-11	COB56D	1818	0.3874 -		SFAM01.1	PPB	8/25/2021 14:41	SMPL
M3460-12	COB56S	11601	2.5662 -		SFAM01.1	PPB	8/25/2021 14:43	SMPL
M3460-13	COB64	2250	0.4836 -		SFAM01.1	PPB	8/25/2021 14:45	SMPL
M3460-14	MCOAB3	1972	0.4217 -		SFAM01.1	PPB	8/25/2021 14:48	SMPL
M3460-15	MCOAB6	2129	0.4567 -		SFAM01.1	PPB	8/25/2021 14:50	SMPL
M3460-16	MCOAS0	1931	0.4126 -		SFAM01.1	PPB	8/25/2021 14:52	SMPL
M3460-17	MCOAS6	1648	0.3495 -		SFAM01.1	PPB	8/25/2021 14:54	SMPL
M3460-18	MCOAR2	757	0.1511 -		SFAM01.1	PPB	8/25/2021 14:57	SMPL
M3460-19	MCOAR8	3140	0.6818 -		SFAM01.1	PPB	8/25/2021 14:59	SMPL
CCV043	CCV043	22626	5.0216 -		SFAM01.1	PPB	8/25/2021 15:01	SMPL
CCB043	CCB043	100	0.0048 -		SFAM01.1	PPB	8/25/2021 15:03	SMPL
M3460-20	MCOAT7	1743	0.3707 -		SFAM01.1	PPB	8/25/2021 15:06	SMPL
M3460-21	MCOAW4	3877	0.846 -		SFAM01.1	PPB	8/25/2021 15:08	SMPL
PB138705BL	PBS705	3	-0.0168 -		SFAM01.1	PPB	8/25/2021 15:14	SMPL
M3473-01	COB89	3541	0.7711 -		SFAM01.1	PPB	8/25/2021 15:17	SMPL
M3473-02	MCOAB9	2147	0.4607 -		SFAM01.1	PPB	8/25/2021 15:19	SMPL
M3473-03	MCOAC0	1948	0.4163 -		SFAM01.1	PPB	8/25/2021 15:21	SMPL
M3473-04	MCOAC1	1433	0.3016 -		SFAM01.1	PPB	8/25/2021 15:23	SMPL
M3473-05	MCOAC2	1589	0.3364 -		SFAM01.1	PPB	8/25/2021 15:26	SMPL
M3473-06	MCOAS3	3070	0.6662 -		SFAM01.1	PPB	8/25/2021 15:28	SMPL
M3473-07	MCOAS3D	2819	0.6103 -		SFAM01.1	PPB	8/25/2021 15:30	SMPL
M3473-08	MCOAS3S	12341	2.731 -		SFAM01.1	PPB	8/25/2021 15:36	SMPL
M3473-09	MCOAX0	599	0.1159 -		SFAM01.1	PPB	8/25/2021 15:38	SMPL
M3473-11	MCOAR6	1689	0.3587 -		SFAM01.1	PPB	8/25/2021 15:40	SMPL
M3473-12	MCOAW2	1586	0.3357 -		SFAM01.1	PPB	8/25/2021 15:42	SMPL

LB116023
INSTRUMENT ID : CV1

M3473-13	MCOAW9	2704	0.5847 -	SFAM01.1	PPB	8/25/2021 15:45	SMPL
M3473-15	MCOAB7	1192	0.248 -	SFAM01.1	PPB	8/25/2021 15:47	SMPL
M3473-16	MCOAB8	920	0.1874 -	SFAM01.1	PPB	8/25/2021 15:49	SMPL
M3473-17	MCOAR5	1911	0.4081 -	SFAM01.1	PPB	8/25/2021 15:51	SMPL
M3473-18	MCOAR9	2582	0.5575 -	SFAM01.1	PPB	8/25/2021 15:54	SMPL
M3473-19	MCOAW0	1175	0.2442 -	SFAM01.1	PPB	8/25/2021 15:56	SMPL
CCV044	CCV044	22375	4.9657 -	SFAM01.1	PPB	8/25/2021 15:58	SMPL
CCB044	CCB044	108	0.0066 -	SFAM01.1	PPB	8/25/2021 16:00	SMPL
M3473-20	MCOAW5	3580	0.7798 -	SFAM01.1	PPB	8/25/2021 16:03	SMPL
PB138706BL	PBS706	123	0.0099 -	SFAM01.1	PPB	8/25/2021 16:05	SMPL
M3492-01	COAY8	622	0.121 -	SFAM01.1	PPB	8/25/2021 16:07	SMPL
M3492-02	COAY9	664	0.1304 -	SFAM01.1	PPB	8/25/2021 16:10	SMPL
M3492-03	COAZ4	1767	0.376 -	SFAM01.1	PPB	8/25/2021 16:17	SMPL
M3492-04	COAZ4D	1818	0.3874 -	SFAM01.1	PPB	8/25/2021 16:19	SMPL
M3492-05	COAZ4S	12122	2.6822 -	SFAM01.1	PPB	8/25/2021 16:21	SMPL
M3492-06	COB05	2182	0.4685 -	SFAM01.1	PPB	8/25/2021 16:23	SMPL
M3492-07	COB19	1051	0.2166 -	SFAM01.1	PPB	8/25/2021 16:26	SMPL
M3492-08	COB24	706	0.1397 -	SFAM01.1	PPB	8/25/2021 16:28	SMPL
M3492-09	COB25	368	0.0645 -	SFAM01.1	PPB	8/25/2021 16:30	SMPL
M3492-10	COB30	1687	0.3582 -	SFAM01.1	PPB	8/25/2021 16:32	SMPL
M3492-11	COB41	1660	0.3522 -	SFAM01.1	PPB	8/25/2021 16:35	SMPL
M3492-12	COB06	1473	0.3106 -	SFAM01.1	PPB	8/25/2021 16:37	SMPL
M3492-13	COB08	2171	0.466 -	SFAM01.1	PPB	8/25/2021 16:39	SMPL
M3492-14	COB42	970	0.1985 -	SFAM01.1	PPB	8/25/2021 16:42	SMPL
M3492-15	MCOAS4	1969	0.421 -	SFAM01.1	PPB	8/25/2021 16:44	SMPL
M3492-16	MCOAX1	2626	0.5673 -	SFAM01.1	PPB	8/25/2021 16:46	SMPL
CCV045	CCV045	22016	4.8858 -	SFAM01.1	PPB	8/25/2021 16:48	SMPL
CCB045	CCB045	99	0.0045 -	SFAM01.1	PPB	8/25/2021 16:51	SMPL