

LB116031


SFAM01.1

INSTRUMENT ID: CV4

Linear

 μ Abs.:

39722



A= 0.0000e+000

B= 2.5054e-004 slope

C= -3.7144e-002 y-intercept

Rho= 0.9998733

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.016	-0.016	85	0.000	85					11
0.05	0.050					573					-5
0.20	0.200	0.191	-0.009	910	0.0 %	910					0
2.50	2.500	2.502	0.002	10133	0.0 %	10133					0
5.00	5.000	4.994	-0.006	20082	0.0 %	20082					2
7.50	7.500	7.614	0.114	30540	0.0 %	30540					1
10.0	10.000	9.915	-0.085	39722	0.0 %	39722					

LB116031
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	85	-		0 SFAM01.1	PPB	8/26/2021 11:52	Std
	0.2 S01	910	-		0.2 SFAM01.1	PPB	8/26/2021 11:54	Std
	2.5 S02	10133	-		2.5 SFAM01.1	PPB	8/26/2021 11:56	Std
	5 S03	20082	-		5 SFAM01.1	PPB	8/26/2021 11:59	Std
	7.5 S04	30540	-		7.5 SFAM01.1	PPB	8/26/2021 12:01	Std
	10 S05	39722	-		10 SFAM01.1	PPB	8/26/2021 12:03	Std
ICV062	ICV062	15420	3.8262 -		SFAM01.1	PPB	8/26/2021 12:06	SMPL
ICB062	ICB062	15	-0.0334 -		SFAM01.1	PPB	8/26/2021 12:08	SMPL
CCV046	CCV046	20120	5.0037 -		SFAM01.1	PPB	8/26/2021 12:10	SMPL
CCB046	CCB046	29	-0.0299 -		SFAM01.1	PPB	8/26/2021 12:13	SMPL
PB138726BL	PBW726	49	-0.0249 -		SFAM01.1	PPB	8/26/2021 12:15	SMPL
M3473-10	MC0AX8	43	-0.0264 -		SFAM01.1	PPB	8/26/2021 12:17	SMPL
M3473-14	MC0AX6	54	-0.0236 -		SFAM01.1	PPB	8/26/2021 12:19	SMPL
PB138727BL	PBW727	62	-0.0216 -		SFAM01.1	PPB	8/26/2021 12:22	SMPL
M3491-01	MC0AC1	73	-0.0189 -		SFAM01.1	PPB	8/26/2021 12:24	SMPL
M3491-02	MC0AC2	163	0.0037 -		SFAM01.1	PPB	8/26/2021 12:26	SMPL
M3491-03	MC0AC4	76	-0.0181 -		SFAM01.1	PPB	8/26/2021 12:28	SMPL
M3491-04	MC0AC5	80	-0.0171 -		SFAM01.1	PPB	8/26/2021 12:31	SMPL
M3491-05	MC0AC6	350	0.0505 -		SFAM01.1	PPB	8/26/2021 12:33	SMPL
M3491-06	MC0AD0	106	-0.0106 -		SFAM01.1	PPB	8/26/2021 12:35	SMPL
M3491-07	MC0AD1	78	-0.0176 -		SFAM01.1	PPB	8/26/2021 12:38	SMPL
M3491-08	MC0AD1D	63	-0.0214 -		SFAM01.1	PPB	8/26/2021 12:40	SMPL
M3491-09	MC0AD1S	4247	1.0269 -		SFAM01.1	PPB	8/26/2021 12:45	SMPL
PB138728BL	PBW728	33	-0.0289 -		SFAM01.1	PPB	8/26/2021 12:47	SMPL
M3492-17	C0B85	48	-0.0251 -		SFAM01.1	PPB	8/26/2021 12:49	SMPL
CCV047	CCV047	19365	4.8146 -		SFAM01.1	PPB	8/26/2021 12:51	SMPL
CCB047	CCB047	9	-0.0349 -		SFAM01.1	PPB	8/26/2021 12:54	SMPL