

LB116032


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

33550



A= 0.0000e+000

B= 2.9658e-004

C= -4.8498e-002

Rho= 0.9998292

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.017	-0.017	107	0.000	107					4
0.2	0.200	0.208	0.008	864	0.0 %	864					-2
2.5	2.500	2.458	-0.042	8451	0.0 %	8451					1
5.0	5.000	5.025	0.025	17106	0.0 %	17106					2
7.5	7.500	7.625	0.125	25872	0.0 %	25872					-1
10.0	10.000	9.902	-0.098	33550	0.0 %	33550					

LB116032 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 26 Aug 2021 14:05:44

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	107	-	0.0000	7470A	PPB	26 Aug 2021 14:10:51	S	Std
0.2 - 1	50.2	864	-	0.2000	7470A	PPB	26 Aug 2021 14:13:07	S	Std
2.5 - 1	52.5	8451	-	2.5000	7470A	PPB	26 Aug 2021 14:15:24	S	Std
5.0 - 1	55	17106	-	5.0000	7470A	PPB	26 Aug 2021 14:17:41	S	Std
7.5 - 1	57.5	25872	-	7.5000	7470A	PPB	26 Aug 2021 14:23:17	S	Std
10.0 - 1	510	33550	-	10.0000	7470A	PPB	26 Aug 2021 14:25:32	S	Std
ICV72 - 1	ICV72	12744	3.7311	-	7470A	PPB	26 Aug 2021 14:28:14	U	SMPL
ICB72 - 1	ICB72	125	-0.0114	-	7470A	PPB	26 Aug 2021 14:30:29	U	SMPL
CCV48 - 1	CCV48	17585	5.1668	-	7470A	PPB	26 Aug 2021 14:32:44	U	SMPL
CCB48 - 1	CCB48	137	-0.0079	-	7470A	PPB	26 Aug 2021 14:34:59	U	SMPL
CRA - 1	CRA	806	0.1905	-	7470A	PPB	26 Aug 2021 14:37:14	U	SMPL
HighStd - 1	HighStd	33592	9.9142	-	7470A	PPB	26 Aug 2021 14:39:28	U	SMPL
ChkStd - 1	ChkStd	23914	7.0439	-	7470A	PPB	26 Aug 2021 14:41:44	U	SMPL
PB138729BL - 1	PBW	13	-0.0446	-	7470A	PPB	26 Aug 2021 14:55:29	U	SMPL
PB138729BS - 1	LCSW	13777	4.0375	-	7470A	PPB	26 Aug 2021 14:57:45	U	SMPL
M3528-02 - 1	PT-3-082421	415	0.0746	-	7470A	PPB	26 Aug 2021 15:02:32	U	SMPL
M3528-02DUP - 1	PT-3-082421DUP	449	0.0847	-	7470A	PPB	26 Aug 2021 15:04:48	U	SMPL
M3528-02MS - 1	PT-3-082421MS	13823	4.0511	-	7470A	PPB	26 Aug 2021 15:07:05	U	SMPL
M3528-02MSD - 1	PT-3-082421MSD	13484	3.9506	-	7470A	PPB	26 Aug 2021 15:11:52	U	SMPL
M3535-02 - 1	WVLY-IDW-SO-082421	2448	0.6775	-	7470A	PPB	26 Aug 2021 15:14:07	U	SMPL
CCV49 - 1	CCV49	17033	5.0031	-	7470A	PPB	26 Aug 2021 15:16:22	U	SMPL
CCB49 - 1	CCB49	131	-0.0096	-	7470A	PPB	26 Aug 2021 15:18:38	U	SMPL
PB138730BL - 1	PBW	13	-0.0446	-	7470A	PPB	26 Aug 2021 15:26:09	U	SMPL
PB138730BS - 1	LCSW	13717	4.0197	-	7470A	PPB	26 Aug 2021 15:28:24	U	SMPL
M3503-01 - 1	RW6A-GW-20210819	2067	0.5645	-	7470A	PPB	26 Aug 2021 15:30:39	U	SMPL
M3506-01 - 1	N29305	1196	0.3062	-	7470A	PPB	26 Aug 2021 15:37:45	U	SMPL
M3506-01DUP - 1	N29305DUP	1151	0.2929	-	7470A	PPB	26 Aug 2021 15:40:01	U	SMPL
M3506-01MS - 1	N29305MS	13254	3.8824	-	7470A	PPB	26 Aug 2021 15:44:49	U	SMPL
M3506-01MSD - 1	N29305MSD	12627	3.6964	-	7470A	PPB	26 Aug 2021 15:47:05	U	SMPL
PB138742BL - 1	PBW	85	-0.0233	-	7470A	PPB	26 Aug 2021 15:54:41	U	SMPL
PB138742BS - 1	LCSW	14178	4.1564	-	7470A	PPB	26 Aug 2021 15:56:58	U	SMPL
PB138729TB - 1	PB138715TB	9	-0.0458	-	7470A	PPB	26 Aug 2021 16:01:59	U	SMPL
CCV50 - 1	CCV50	17706	5.2027	-	7470A	PPB	26 Aug 2021 16:04:15	U	SMPL
CCB50 - 1	CCB50	129	-0.0102	-	7470A	PPB	26 Aug 2021 16:06:31	U	SMPL
PB138742TB - 1	PB138587TB	36	-0.0378	-	7470A	PPB	26 Aug 2021 16:08:47	U	SMPL
M3471-15 - 1	2021-08-17-EB	198	0.0102	-	7470A	PPB	26 Aug 2021 16:13:48	U	SMPL
M3528-02LX5 - 1		89	-0.0221	-	7470A	PPB	26 Aug 2021 16:16:02	U	SMPL
M3528-02A - 1		13033	3.8168	-	7470A	PPB	26 Aug 2021 16:18:18	U	SMPL
M3506-01LX5 - 1		170	0.0019	-	7470A	PPB	26 Aug 2021 16:20:33	U	SMPL
M3506-01A - 1		13574	3.9773	-	7470A	PPB	26 Aug 2021 16:22:50	U	SMPL
CCV51 - 1	CCV51	16953	4.9794	-	7470A	PPB	26 Aug 2021 16:25:06	U	SMPL
CCB51 - 1	CCB51	-1	-0.0488	-	7470A	PPB	26 Aug 2021 16:27:23	U	SMPL