



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB116036

Slope : 98.1

BalanceID: Wc Sc-06

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	22.1	4.01	08/26/2021	17:10
2	CAL2	1	Water	NA	NA	22.1	7.00	08/26/2021	17:11
3	CAL3	1	Water	NA	NA	22.1	10.03	08/26/2021	17:12
4	ICV	1	Water	NA	NA	22.2	6.99	08/26/2021	17:15
5	CCV1	1	Water	NA	NA	22.2	2.01	08/26/2021	17:16
6	M3550-01	1	Solid	20.03	20	24.8	7.02	08/26/2021	17:19
7	M3550-01DUP	1	Solid	20.04	20	24.9	7.03	08/26/2021	17:20
8	CCV2	1	Water	NA	NA	22.2	12.02	08/26/2021	17:22

VB 116036

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity m3550 WorkList ID : 151792 Department : Wet-Chemistry Date : 08-26-2021 16:59:30

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/26/2021	Solid	M3550-01	☒ Corrosivity	Cool 4 deg C	WOOD08		WC-1	08/26/2021	9045D

Date/Time 08/26/21 17:00
Raw Sample Received by: 90 (WC)
Raw Sample Relinquished by: KS (WC)

Date/Time 08/26/21 18:00
Raw Sample Received by: KS (WC)
Raw Sample Relinquished by: 90 (WC)