

LB 116039


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

53885



A= 0.0000e+000

B= 1.8803e-004

C= 2.3914e-002

Rho= 0.9993739

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.048	0.048	126	0.000	126					-3
0.2	0.200	0.207	0.007	973	0.0 %	973					-1
2.5	2.500	2.480	-0.020	13064	0.0 %	13064					1
5.0	5.000	5.073	0.073	26854	0.0 %	26854					-4
7.5	7.500	7.236	-0.264	38354	0.0 %	38354					2
10.0	10.000	10.156	0.156	53885	0.0 %	53885					

LB116039 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 27 Aug 2021 08:42:38

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	126	-	0.0000	7471B	PPB	27 Aug 2021 08:48:32	S	Std
0.2 - 1	50.2	973	-	0.2000	7471B	PPB	27 Aug 2021 08:50:49	S	Std
2.5 - 1	52.5	13064	-	2.5000	7471B	PPB	27 Aug 2021 08:53:06	S	Std
5.0 - 1	55	26854	-	5.0000	7471B	PPB	27 Aug 2021 08:55:23	S	Std
7.5 - 1	57.5	38354	-	7.5000	7471B	PPB	27 Aug 2021 08:57:40	S	Std
10.0 - 1	510	53885	-	10.0000	7471B	PPB	27 Aug 2021 08:59:56	S	Std
ICV73 - 1	ICV73	19513	3.6930	-7471B	PPB	PPB	27 Aug 2021 09:02:45	U	SMPL
ICB73 - 1	ICB73	-28	0.0186	-7471B	PPB	PPB	27 Aug 2021 09:04:59	U	SMPL
CCV52 - 1	CCV52	26881	5.0784	-7471B	PPB	PPB	27 Aug 2021 09:07:14	U	SMPL
CCB52 - 1	CCB52	-29	0.0185	-7471B	PPB	PPB	27 Aug 2021 09:09:29	U	SMPL
CRA - 1	CRA	949	0.2024	-7471B	PPB	PPB	27 Aug 2021 09:11:44	U	SMPL
HighStd - 1	HighStd	53428	10.0701	-7471B	PPB	PPB	27 Aug 2021 09:13:59	U	SMPL
ChkStd - 1	ChkStd	36429	6.8738	-7471B	PPB	PPB	27 Aug 2021 09:16:14	U	SMPL
PB138725BL - 1	PBS	-100	0.0051	-7471B	PPB	PPB	27 Aug 2021 09:18:31	U	SMPL
PB138725BS - 1	LCS	18260	3.4574	-7471B	PPB	PPB	27 Aug 2021 09:20:48	U	SMPL
M3528-01 - 1	PT-3-082421	80644	15.1876	-7471B	PPB	PPB	27 Aug 2021 09:23:04	U	SMPL
M3530-01 - 1	PL-01-082421	1328	0.2736	-7471B	PPB	PPB	27 Aug 2021 09:25:20	U	SMPL
M3533-01 - 1	CL-02-0824-21	1875	0.3765	-7471B	PPB	PPB	27 Aug 2021 09:27:39	U	SMPL
M3538-01 - 1	VNJ-210	4675	0.9030	-7471B	PPB	PPB	27 Aug 2021 09:36:04	U	SMPL
M3538-01DUP - 1	VNJ-210DUP	4984	0.9611	-7471B	PPB	PPB	27 Aug 2021 09:38:19	U	SMPL
CCV53 - 1	CCV53	28832	5.4453	-7471B	PPB	PPB	27 Aug 2021 09:40:34	U	SMPL
CCB53 - 1	CCB53	-27	0.0188	-7471B	PPB	PPB	27 Aug 2021 09:42:49	U	SMPL
M3538-01MS - 1	VNJ-210MS	23120	4.3712	-7471B	PPB	PPB	27 Aug 2021 09:45:05	U	SMPL
M3538-01MSD - 1	VNJ-210MSD	23319	4.4087	-7471B	PPB	PPB	27 Aug 2021 09:47:20	U	SMPL
M3540-01 - 1	OR-2-082521	878	0.1890	-7471B	PPB	PPB	27 Aug 2021 09:49:35	U	SMPL
M3543-01 - 1	SP	6172	1.1845	-7471B	PPB	PPB	27 Aug 2021 09:51:51	U	SMPL
M3545-01 - 1	NCS-COMP-1	9375	1.7867	-7471B	PPB	PPB	27 Aug 2021 09:54:06	U	SMPL
M3545-03 - 1	NCS-COMP-2	364632	HIGH	-7471B	PPB	PPB	27 Aug 2021 09:56:23	X	
M3546-01 - 1	TPR-4-COMP	5594	1.0758	-7471B	PPB	PPB	27 Aug 2021 09:58:12	U	SMPL
M3546-03 - 1	TRP-7-COMP	3297	0.6439	-7471B	PPB	PPB	27 Aug 2021 10:00:31	U	SMPL
M3548-01 - 1	VNJ-235	10875	2.0688	-7471B	PPB	PPB	27 Aug 2021 10:02:48	U	SMPL
M3557-01 - 1	OK-02-082621	4707	0.9090	-7471B	PPB	PPB	27 Aug 2021 10:05:04	U	SMPL
CCV54 - 1	CCV54	28933	5.4643	-7471B	PPB	PPB	27 Aug 2021 10:07:19	U	SMPL
CCB54 - 1	CCB54	14	0.0265	-7471B	PPB	PPB	27 Aug 2021 10:09:34	U	SMPL
M3558-01 - 1	TPR-14-COMP	2379	0.4712	-7471B	PPB	PPB	27 Aug 2021 10:11:50	U	SMPL
M3545-03DLX10 - 1		57497	10.8352	-7471B	PPB	PPB	27 Aug 2021 10:14:05	U	SMPL
M3538-01LX5 - 1		751	0.1651	-7471B	PPB	PPB	27 Aug 2021 10:16:20	U	SMPL
M3538-01A - 1		26886	5.0794	-7471B	PPB	PPB	27 Aug 2021 10:18:37	U	SMPL
M3528-01DLX5 - 1		17912	3.3920	-7471B	PPB	PPB	27 Aug 2021 10:20:53	U	SMPL
M3545-03DLX25 - 1		22993	4.3474	-7471B	PPB	PPB	27 Aug 2021 10:23:09	U	SMPL
CCV55 - 1	CCV55	25415	4.8028	-7471B	PPB	PPB	27 Aug 2021 10:25:25	U	SMPL
CCB55 - 1	CCB55	-16	0.0209	-7471B	PPB	PPB	27 Aug 2021 10:27:42	U	SMPL