

LB 116040

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : AB Instrument ID : Konelab

8/27/2021 11:09

Test: ^{Positive} Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	96.318	0.0	0.078	
ICB1	1.290	0.0	0.001	
CCV1	235.169	0.0	0.190	
CCB1	1.303	0.0	0.001	
PB138745BL	1.113	0.0	0.001	
M3528-02	1.142	0.0	0.001	
M3528-02DUP	1.150	0.0	0.001	
M3550-01	1.365	0.0	0.001	
CCV2	241.674	0.0	0.195	
CCB2	1.162	0.0	0.001	

N	10
Mean	58.169
SD	99.5335
CV%	171.11

Aquakem v. 7.2AQ1

Results from time period:

Fri Aug 27 08:44:30 2021

Fri Aug 27 11:05:22 2021

Sample Id	Sam/Ctr/c/	Test short nar	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	0.8005	µg/l	8/27/2021 9:41:19	
5.OPPBCN	A	Reactive CN	P	5.553	µg/l	8/27/2021 9:41:20	
10PPBCN	A	Reactive CN	P	10.2776	µg/l	8/27/2021 9:41:21	
50PPBCN	A	Reactive CN	P	50.6478	µg/l	8/27/2021 9:41:22	
100PPBCN	A	Reactive CN	P	98.8288	µg/l	8/27/2021 9:41:23	
250PPBCN	A	Reactive CN	P	247.468	µg/l	8/27/2021 9:41:24	
500PPBCN	A	Reactive CN	P	501.4244	µg/l	8/27/2021 9:41:25	
ICV1	S	Reactive CN	P	96.3182	µg/l	8/27/2021 10:49:07	
ICB1	S	Reactive CN	P	1.2903	µg/l	8/27/2021 10:49:09	
CCV1	S	Reactive CN	P	235.1686	µg/l	8/27/2021 10:49:12	
CCB1	S	Reactive CN	P	1.303	µg/l	8/27/2021 10:49:13	
PB138745BL	S	Reactive CN	P	1.1128	µg/l	8/27/2021 10:49:16	
M3528-02	S	Reactive CN	P	1.1424	µg/l	8/27/2021 10:56:42	
M3528-02DUP	S	Reactive CN	P	1.1501	µg/l	8/27/2021 10:56:44	
M3550-01	S	Reactive CN	P	1.365	µg/l	8/27/2021 10:56:45	
CCV2	S	Reactive CN	P	241.6738	µg/l	8/27/2021 10:56:47	
CCB2	S	Reactive CN	P	1.1622	µg/l	8/27/2021 10:56:50	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : AB Instrument ID : Konelab

8/27/2021 9:41


Test Total CN

Accepted 8/27/2021 9:41

Factor 1237
Bias 0

Coeff. of det. 0.999945

Errors



	Calibrator	Response	Calc. con.	Conc.	Rel. Errors
1	0.0PPBCN	0.001	0.8005	0.0000	
2	5.0PPBCN	0.005	5.5530	5.0000	11.1
3	10PPBCN	0.008	10.2776	10.0000	2.8
4	50PPBCN	0.041	50.6478	50.0000	1.3
5	100PPBCN	0.080	98.8288	100.0000	1.2
6	250PPBCN	0.200	247.4680	250.0000	-1.0
7	500PPBCN	0.405	501.4244	500.0000	0.3

AB
8/27/21