

LB116041


245.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

37901



A= 0.0000e+000

B= 2.6768e-004

C= -3.7097e-002

Rho= 0.9995877

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.012	-0.012	92	0.000	92					
0.05	0.050	0.065	0.015	381	0.0 %	381	0				30
0.2	0.200	0.199	-0.001	882	0.0 %	882					-1
2.5	2.500	2.512	0.012	9522	0.0 %	9522					0
5.0	5.000	5.112	0.112	19238	0.0 %	19238					2
7.5	7.500	7.266	-0.234	27284	0.0 %	27284					-3
10.0	10.000	10.108	0.108	37901	0.0 %	37901					1

LB116041 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 27 Aug 2021 12:14:59

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	92	-	0.0000	245.1	PPB	27 Aug 2021 12:24:36	S	Std
0.05 - 1	50.05	381	-	0.0500	245.1	PPB	27 Aug 2021 12:26:53	S	Std
0.2 - 1	50.2	882	-	0.2000	245.1	PPB	27 Aug 2021 12:29:09	S	Std
2.5 - 1	54.5	9522	-	2.5000	245.1	PPB	27 Aug 2021 12:31:26	S	Std
5.0 - 1	55	19238	-	5.0000	245.1	PPB	27 Aug 2021 12:33:43	S	Std
7.5 - 1	57.5	27284	-	7.5000	245.1	PPB	27 Aug 2021 12:36:00	S	Std
10.0 - 1	510	37901	-	10.0000	245.1	PPB	27 Aug 2021 12:48:59	S	Std
ICV74 - 1	ICV74	14630	3.8790	-245.1		PPB	27 Aug 2021 12:52:00	U	SMPL
ICB74 - 1	ICB74	15	-0.0331	-245.1		PPB	27 Aug 2021 12:54:14	U	SMPL
CCV56 - 1	CCV56	18906	5.0236	-245.1		PPB	27 Aug 2021 12:56:29	U	SMPL
CCB56 - 1	CCB56	9	-0.0347	-245.1		PPB	27 Aug 2021 12:58:44	U	SMPL
CRA - 1	CRA	366	0.0609	-245.1		PPB	27 Aug 2021 13:00:59	U	SMPL
HighStd - 1	HighStd	37573	10.0204	-245.1		PPB	27 Aug 2021 13:03:14	U	SMPL
ChkStd - 1	ChkStd	26168	6.9675	-245.1		PPB	27 Aug 2021 13:05:29	U	SMPL
PB138765BL - 1	PBW	237	0.0263	-245.1		PPB	27 Aug 2021 13:07:47	U	SMPL
PB138765BS - 1	LCSW	14313	3.7942	-245.1		PPB	27 Aug 2021 13:10:04	U	SMPL
M3481-01 - 1	20211007-DI-WATER	230	0.0245	-245.1		PPB	27 Aug 2021 13:12:21	U	SMPL
M3493-01 - 1	20211008-MANASQUAN-WATER	439	0.0804	-245.1		PPB	27 Aug 2021 13:14:37	U	SMPL
M3493-01DUP - 1	20211008-MANASQUAN-WATERDUP	406	0.0716	-245.1		PPB	27 Aug 2021 13:16:54	U	SMPL
M3493-01MS - 1	20211008-MANASQUAN-WATERMS	17534	4.6564	-245.1		PPB	27 Aug 2021 13:19:11	U	SMPL
M3493-01MSD - 1	20211008-MANASQUAN-WATERMSD	16456	4.3678	-245.1		PPB	27 Aug 2021 13:21:26	U	SMPL
CCV57 - 1	CCV57	18860	5.0113	-245.1		PPB	27 Aug 2021 13:23:42	U	SMPL
CCB57 - 1	CCB57	-8	-0.0392	-245.1		PPB	27 Aug 2021 13:25:58	U	SMPL
M3493-01LX5 - 1		66	-0.0194	-245.1		PPB	27 Aug 2021 13:28:13	U	SMPL
M3493-01A - 1		16825	4.4666	-245.1		PPB	27 Aug 2021 13:30:29	U	SMPL
CCV58 - 1	CCV58	19137	5.0855	-245.1		PPB	27 Aug 2021 13:32:44	U	SMPL
CCB58 - 1	CCB58	0	-0.0371	-245.1		PPB	27 Aug 2021 13:34:59	U	SMPL