

LB116048


7470A

INSTRUMENT ID: CV2

Linear

 μ Abs.:

40063



A= 0.0000e+000

B= 2.5426e-004

C= -5.1162e-002

Rho= 0.9994324

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.009	0.009	235	0.000	235					
0.2	0.200	0.173	-0.027	883	0.0 %	883					-14
2.5	2.500	2.590	0.090	10386	0.0 %	10386					4
5.0	5.000	5.045	0.045	20044	0.0 %	20044					1
7.5	7.500	7.248	-0.252	28707	0.0 %	28707					-3
10.0	10.000	10.135	0.135	40063	0.0 %	40063					1

LB116048 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 27 Aug 2021 15:18:17

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	235	-	0.0000	7470A	PPB	27 Aug 2021 15:21:12	S	Std
0.2 - 1	50.2	883	-	0.2000	7470A	PPB	27 Aug 2021 15:23:29	S	Std
2.5 - 1	52.5	10386	-	2.5000	7470A	PPB	27 Aug 2021 15:25:46	S	Std
5.0 - 1	55	20044	-	5.0000	7470A	PPB	27 Aug 2021 15:28:02	S	Std
7.5 - 1	57.5	28707	-	7.5000	7470A	PPB	27 Aug 2021 15:30:19	S	Std
10.0 - 1	50	40063	-	10.0000	7470A	PPB	27 Aug 2021 15:35:13	S	Std
ICV75 - 1	ICV75	14915	3.7411	-	7470A	PPB	27 Aug 2021 15:40:08	U	SMPL
ICB75 - 1	ICB75	166	-0.0090	-	7470A	PPB	27 Aug 2021 15:42:22	U	SMPL
CCV59 - 1	CCV59	19806	4.9847	-	7470A	PPB	27 Aug 2021 15:44:37	U	SMPL
CCB59 - 1	CCB59	163	-0.0097	-	7470A	PPB	27 Aug 2021 15:46:52	U	SMPL
CRA - 1	CRA	846	0.1639	-	7470A	PPB	27 Aug 2021 15:51:59	U	SMPL
HighStd - 1	HighStd	40664	10.2881	-	7470A	PPB	27 Aug 2021 15:54:13	U	SMPL
ChkStd - 1	ChkStd	25757	6.4978	-	7470A	PPB	27 Aug 2021 15:56:28	U	SMPL
PB138759BL - 1	PBW	-13	-0.0545	-	7470A	PPB	27 Aug 2021 15:58:44	U	SMPL
PB138759BS - 1	LCSW	14842	3.7226	-	7470A	PPB	27 Aug 2021 16:01:00	U	SMPL
M3550-01 - 1	WC-1	35	-0.0423	-	7470A	PPB	27 Aug 2021 16:03:16	U	SMPL
M3550-01DUP - 1	WC-1DUP	67	-0.0341	-	7470A	PPB	27 Aug 2021 16:05:33	U	SMPL
M3550-01MS - 1	WC-1MS	14611	3.6638	-	7470A	PPB	27 Aug 2021 16:07:49	U	SMPL
M3550-01MSD - 1	WC-1MSD	14944	3.7485	-	7470A	PPB	27 Aug 2021 16:10:06	U	SMPL
CCV60 - 1	CCV60	19982	5.0295	-	7470A	PPB	27 Aug 2021 16:12:21	U	SMPL
CCB60 - 1	CCB60	149	-0.0133	-	7470A	PPB	27 Aug 2021 16:14:37	U	SMPL
PB138759TB - 1	PB138746TB	86	-0.0293	-	7470A	PPB	27 Aug 2021 16:16:52	U	SMPL
M3550-01LX5 - 1		190	-0.0029	-	7470A	PPB	27 Aug 2021 16:19:07	U	SMPL
M3550-01A - 1		15535	3.8988	-	7470A	PPB	27 Aug 2021 16:23:19	U	SMPL
CCV61 - 1	CCV61	19742	4.9685	-	7470A	PPB	27 Aug 2021 16:25:34	U	SMPL
CCB61 - 1	CCB61	171	-0.0077	-	7470A	PPB	27 Aug 2021 16:27:50	U	SMPL