



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB116051

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	08/27/2021	15:20
2	CAL2	1	Water	NA	NA	21.3	7.00	08/27/2021	15:21
3	CAL3	1	Water	NA	NA	21.3	10.03	08/27/2021	15:22
4	ICV	1	Water	NA	NA	21.5	6.99	08/27/2021	15:25
5	CCV1	1	Water	NA	NA	21.4	2.01	08/27/2021	15:30
6	M3582-01	1	Water	NA	NA	26.1	5.03	08/27/2021	15:32
7	M3582-01DUP	1	Water	NA	NA	26.2	5.04	08/27/2021	15:33
8	CCV2	1	Water	NA	NA	21.4	12.02	08/27/2021	15:35

2B 116051

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph m3582 WorkList ID : 151819 Department : Wet-Chemistry Date : 08-27-2021 15:10:01

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/26/2021	Water	M3582-01	A pH	Cool 4 deg C	STER08	D11	POOL-WATER	08/26/2021	9040C

Date/Time 08/27/21 15:15
Raw Sample Received by: JP (WPC)
Raw Sample Relinquished by: CL SM

Date/Time 08/27/21 16:10
Raw Sample Received by: CL SM
Raw Sample Relinquished by: JP (WPC)