

LB116063


SFAM01.1

INSTRUMENT ID: CV2

Linear

 μ Abs.:

41729



A= 0.0000e+000

B= 2.4000e-004 *slope*C= -8.0451e-004 *intercept*

Rho= 0.9999805

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.023	0.023	100	0.000	100					-
0.05	0.050					573					8
0.20	0.200	0.215	0.015	898	0.0 %	898					-1
2.50	2.500	2.467	-0.033	10284	0.0 %	10284					-1
5.00	5.000	4.968	-0.032	20704	0.0 %	20704					0
7.50	7.500	7.512	0.012	31304	0.0 %	31304					0
10.0	10.000	10.014	0.014	41729	0.0 %	41729					0

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	100	-		0 SFAM01.1	PPB	8/30/2021 13:44	Std
	0.2 S01	898	-		0.2 SFAM01.1	PPB	8/30/2021 13:46	Std
	2.5 S02	10284	-		2.5 SFAM01.1	PPB	8/30/2021 13:48	Std
	5 S03	20704	-		5 SFAM01.1	PPB	8/30/2021 13:50	Std
	7.5 S04	31304	-		7.5 SFAM01.1	PPB	8/30/2021 13:53	Std
	10 S05	41729	-		10 SFAM01.1	PPB	8/30/2021 13:58	Std
ICV063	ICV063	16597	3.9825 -		SFAM01.1	PPB	8/30/2021 14:00	SMPL
ICB063	ICB063	-11	-0.0034 -		SFAM01.1	PPB	8/30/2021 14:02	SMPL
CCV062	CCV062	20846	5.0023 -		SFAM01.1	PPB	8/30/2021 14:05	SMPL
CCB062	CCB062	-12	-0.0037 -		SFAM01.1	PPB	8/30/2021 14:07	SMPL
PB138788BL	PBW788	36	0.0078 -		SFAM01.1	PPB	8/30/2021 14:09	SMPL
PB138788TB	LEB788	71	0.0162 -		SFAM01.1	PPB	8/30/2021 14:12	SMPL
M3555-01	MCOAH0	60	0.0136 -		SFAM01.1	PPB	8/30/2021 14:14	SMPL
M3555-02	MCOAH1	74	0.017 -		SFAM01.1	PPB	8/30/2021 14:16	SMPL
M3555-03	MCOAH2	73	0.0167 -		SFAM01.1	PPB	8/30/2021 14:18	SMPL
M3555-04	MCOAH3	113	0.0263 -		SFAM01.1	PPB	8/30/2021 14:21	SMPL
M3555-05	MCOAH3D	105	0.0244 -		SFAM01.1	PPB	8/30/2021 14:23	SMPL
M3555-06	MCOAH3S	5027	1.2057 -		SFAM01.1	PPB	8/30/2021 14:25	SMPL
PB138789BL	PBW789	57	0.0129 -		SFAM01.1	PPB	8/30/2021 14:27	SMPL
PB138789TB	LEB789	70	0.016 -		SFAM01.1	PPB	8/30/2021 14:30	SMPL
M3556-01	MCOAH4	86	0.0198 -		SFAM01.1	PPB	8/30/2021 14:32	SMPL
M3556-02	MCOAH5	81	0.0186 -		SFAM01.1	PPB	8/30/2021 14:34	SMPL
M3556-03	MCOAH6	71	0.0162 -		SFAM01.1	PPB	8/30/2021 14:36	SMPL
M3556-04	MCOAH7	72	0.0165 -		SFAM01.1	PPB	8/30/2021 14:39	SMPL
M3556-05	MCOAH7D	70	0.016 -		SFAM01.1	PPB	8/30/2021 14:41	SMPL
M3556-06	MCOAH7S	4899	1.175 -		SFAM01.1	PPB	8/30/2021 14:43	SMPL
CCV063	CCV063	20925	5.0212 -		SFAM01.1	PPB	8/30/2021 14:45	SMPL
CCB063	CCB063	18	0.0035 -		SFAM01.1	PPB	8/30/2021 14:48	SMPL