

LB116064


SFAM01.1

INSTRUMENT ID: CV2

Linear

 μ Abs.:

41379



A= 0.0000e+000

B= 2.4279e-004 *slope*C= 1.4297e-002 *y-intercept*

Rho= 0.9999391

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>901)</i>
0.00	0.000	0.030	0.030	66	0.000	66					=
0.05	0.050					573					10
0.20	0.200	0.219	0.019	845	0.0 %	845					11
2.50	2.500	2.486	-0.014	10182	0.0 %	10182					11
5.00	5.000	4.944	-0.056	20305	0.0 %	20305					11
7.50	7.500	7.459	-0.041	30664	0.0 %	30664					11
10.0	10.000	10.061	0.061	41379	0.0 %	41379					)

LB116064

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type
	0 S0	66	-		0 SFAM01.1	PPB	8/30/2021 15:04	Std
	0.2 S01	845	-		0.2 SFAM01.1	PPB	8/30/2021 15:07	Std
	2.5 S02	10182	-		2.5 SFAM01.1	PPB	8/30/2021 15:09	Std
	5 S03	20305	-		5 SFAM01.1	PPB	8/30/2021 15:11	Std
	7.5 S04	30664	-		7.5 SFAM01.1	PPB	8/30/2021 15:13	Std
	10 S05	41379	-		10 SFAM01.1	PPB	8/30/2021 15:16	Std
ICV064	ICV064	16025	3.905 -		SFAM01.1	PPB	8/30/2021 15:18	SMPL
ICB064	ICB064	-3	0.0136 -		SFAM01.1	PPB	8/30/2021 15:21	SMPL
CCV064	CCV064	20398	4.9667 -		SFAM01.1	PPB	8/30/2021 15:23	SMPL
CCB064	CCB064	-17	0.0102 -		SFAM01.1	PPB	8/30/2021 15:25	SMPL
PB138802BL	PBW802	39	0.0238 -		SFAM01.1	PPB	8/30/2021 15:27	SMPL
M3516-01	MCOAB7	60	0.0289 -		SFAM01.1	PPB	8/30/2021 15:30	SMPL
M3516-02	MCOAB8	80	0.0337 -		SFAM01.1	PPB	8/30/2021 15:32	SMPL
M3516-03	MCOAB9	70	0.0313 -		SFAM01.1	PPB	8/30/2021 15:34	SMPL
M3516-04	MCOAB9D	63	0.0296 -		SFAM01.1	PPB	8/30/2021 15:36	SMPL
M3516-05	MCOAB9S	5044	1.2389 -		SFAM01.1	PPB	8/30/2021 15:39	SMPL
M3516-06	MCOAC0	48	0.026 -		SFAM01.1	PPB	8/30/2021 15:41	SMPL
M3516-07	MCOAC9	187	0.0597 -		SFAM01.1	PPB	8/30/2021 15:43	SMPL
CCV065	CCV065	20511	4.9941 -		SFAM01.1	PPB	8/30/2021 15:46	SMPL
CCB065	CCB065	-28	0.0075 -		SFAM01.1	PPB	8/30/2021 15:48	SMPL