



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB116078

Slope : 98.7

BalanceID: Wc Sc-6

pH Meter ID : WC PH MBTER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	22.1	4.01	08/30/2021	13:35
2	CAL2	1	Water	NA	NA	22.1	7.00	08/30/2021	13:36
3	CAL3	1	Water	NA	NA	22.2	10.03	08/30/2021	13:37
4	ICV	1	Water	NA	NA	22.4	6.99	08/30/2021	13:39
5	CCV1	1	Water	NA	NA	22.5	2.01	08/30/2021	13:40
6	M3593-02	1	Solid	20.03	20	23.6	8.00	08/30/2021	13:42
7	M3593-02DUP	1	Solid	20.04	20	23.7	8.01	08/30/2021	13:43
8	CCV2	1	Water	NA	NA	23.3	12.02	08/30/2021	13:45

116078

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY m3593 WorkList ID : 151853 Department : Wet-Chemistry Date : 08-30-2021 13:31:08

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/30/2021	Solid	M3593-02	Corrosivity	Cool 4 deg C	PSEG03	H11	ETGI-280	08/30/2021	9045D

Date/Time 08/30/21 13:30
Raw Sample Received by: JP LWC
Raw Sample Relinquished by: JP LWC

Date/Time 08/30/21 14:30
Raw Sample Received by: JP LWC
Raw Sample Relinquished by: JP LWC