



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB116079

Slope : 98.9

pH Meter ID : WC PH MBTER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	22.5	4.01	08/30/2021	17:20
2	CAL2	1	Water	NA	NA	22.5	7.00	08/30/2021	17:21
3	CAL3	1	Water	NA	NA	22.5	10.03	08/30/2021	17:22
4	ICV	1	Water	NA	NA	22.6	7.01	08/30/2021	17:24
5	CCV1	1	Water	NA	NA	22.5	2.01	08/30/2021	17:25
6	M3603-01	1	Water	NA	NA	21.3	7.87	08/30/2021	17:29
7	M3603-01DUP	1	Water	NA	NA	21.4	7.88	08/30/2021	17:30
8	M3603-05	1	Water	NA	NA	23.0	6.90	08/30/2021	17:33
9	M3603-06	1	Water	NA	NA	23.1	6.47	08/30/2021	17:36
10	CCV2	1	Water	NA	NA	22.5	12.02	08/30/2021	17:39

WORKLIST(Hardcopy Internal Chain)

VB 116079

WorkList Name : PH M3603

WorkList ID : 151880

Department : Wet-Chemistry

Date : 08-30-2021 17:12:13

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/30/2021	Water	M3603-01	pH	Cool 4 deg C	PSEG03	H11	TRE-1242	08/30/2021	9040C
08/30/2021	Water	M3603-05	pH	Cool 4 deg C	PSEG03	H11	TRE-1243	08/30/2021	9040C
08/30/2021	Water	M3603-06	pH	Cool 4 deg C	PSEG03	H11	TRE-1244	08/30/2021	9040C

Date/Time 08/30/21 17:15

Raw Sample Received by: JB (COC)

Raw Sample Relinquished by: JB (COC)

Date/Time 08/30/21 18:00

Raw Sample Received by: JB (COC)

Raw Sample Relinquished by: JB (COC)