

Analytical Summary ReportAnalysis Method: SM4500 Cl GANALYST: apatelParameter: Residual ChlorineSUPERVISOR REVIEW BY: ketankumarRun Number: LB116080

Reagent/Standard	Lot/Log #
Residual chlorine LOD, 0.05PPM	WP92742
Residual chlorine ICV-LCS, 0.4PPM	WP92741
Chlorine Calibration std, 0.1ppm	WP92736
Chlorine Calibration std, 0.2ppm	WP92737
Chlorine Calibration std, 0.8ppm	WP92739
Chlorine Calibration std, 0.0ppm	WP92735
Chlorine Calibration std, 1.6ppm	WP92740
Residual Chlorine Calibration and CCV std, 0	WP92738
Total Chlorine Powder Pillows	W2759

Intercept: 0.0054Slope: 0.9540Regression: 0.999982

Seq	Lab ID	True Val (mg/l)	DF	Initial Reading	Final Reading	Difference	Result (mg/l)	%D	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	-0.01		08/31/2021	09:08
2	CAL2	0.1	1	0.000	0.100	0.100	0.10	1	08/31/2021	09:14
3	CAL3	0.2	1	0.000	0.200	0.200	0.20	-2	08/31/2021	09:20
4	CAL4	0.4	1	0.000	0.390	0.390	0.40	-0.8	08/31/2021	09:25
5	CAL5	0.8	1	0.000	0.770	0.770	0.80	-0.1	08/31/2021	09:31
6	CAL6	1.6	1	0.000	1.530	1.530	1.60	0.1	08/31/2021	09:36



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: apatel

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: ketankumar

Run Number: LB116080

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.4100	0.4100	0.4240	08/31/2021	09:42
2	ICB				1	0.0000	0.0000	0.0000	-0.0060	08/31/2021	09:48
3	CCV1			0.4	1	0.0000	0.4100	0.4100	0.4240	08/31/2021	09:55
4	CCB1				1	0.0000	0.0000	0.0000	-0.0060	08/31/2021	10:01
5	LB116080BL	50	50		1	0.0000	0.0000	0.0000	-0.0060	08/31/2021	10:08
6	LB116080BS	50	50	0.4	1	0.0000	0.4100	0.4100	0.4240	08/31/2021	10:14
7	M2940-09	50	50		1	0.0000	0.0600	0.0600	0.0570	08/31/2021	10:20
8	CCV2			0.4	1	0.0000	0.4000	0.4000	0.4140	08/31/2021	10:26
9	CCB2				1	0.0000	0.0000	0.0000	-0.0060	08/31/2021	10:35