



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB116083

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	08/31/2021	10:40
2	M3603-01		1	12	100.00	>212.0	>212.0	08/31/2021	11:05
3	M3603-01DUP		1	12	100.00	>212.0	>212.0	08/31/2021	11:25
4	M3603-03		1	10	100.00	>212.0	>212.0	08/31/2021	11:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB116083

WorkList Name : FP83121 WorkList ID : 151897 Department : Wet-Chemistry Date : 08-31-2021 09:11:57

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/02/2021	Water	M3603-01	Flash Point	Cool 4 deg C	PSEG03	H11	TRE-1242	08/30/2021	1010B
09/02/2021	Water	M3603-03	Flash Point	Cool 4 deg C	PSEG03	H11	TRE-1243-TRE-1244	08/30/2021	1010B

Date/Time 08/31/21 09:50
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: RJ CEXT-1061

Date/Time 08/31/21 11:55
 Raw Sample Received by: RJ CEXT-1061
 Raw Sample Relinquished by: KS (WC)