



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatelANALYST: jigneshDate: 09/01/2021OVEN IN TEMP1: 104 °COVEN OUT TEMP1: 103 °CDATE & TIME IN1: 08/31/2021 16:00 DATE & TIME OUT1: 09/01/2021 08:00OVEN IN TEMP2: 104 °COVEN OUT TEMP2: 104 °CDATE & TIME IN2: 09/01/2021 08:30 DATE & TIME OUT2: 09/01/2021 10:00Run Number: LB116090BalanceID: Wc Sc-6OvenID: Wc Oven-1

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB116090BL	LB116090BL	85.3671	85.3671	100	85.3672	85.3672	85.3672	0.0001	1
2	M3610-01	YORK-ST-MH-1	90.0026	90.0026	100	90.2392	90.2392	90.2392	0.2366	2366
3	M3610-01DUP	YORK-ST-MH-1DUP	90.0382	90.0382	100	90.2693	90.2693	90.2693	0.2311	2311
4	M3610-02	YORK-ST-MH-2	90.7332	90.7332	100	90.7508	90.7508	90.7508	0.0176	176

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result mg/L} = ((C - B) / A) * 1000 * 1000$$

116090

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS M3610 WorkList ID : 151908 Department : Wet-Chemistry Date : 08-31-2021 12:11:54

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/03/2021	Water	M3610-01	TS	Cool 4 deg C	DIST01	D11	YORK-ST-MH-1	08/31/2021	SM2540 B
09/03/2021	Water	M3610-02	TS	Cool 4 deg C	DIST01	D11	YORK-ST-MH-2	08/31/2021	SM2540 B

Date/Time 08/31/21 12:30
Raw Sample Received by: JP (WC)
Raw Sample Relinquished by: CF SM

Date/Time 08/31/21 16:10
Raw Sample Received by: CF SM
Raw Sample Relinquished by: JP (WC)