



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 9/1/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 16:50
In Date: 08/31/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:10
Out Date: 09/01/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB116094

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
M3605-01	PX-1	1	1.15	9.69	8.63	87.6	
M3605-02	PX-2	2	1.13	9.9	9.08	90.6	
M3605-03	PX-3	3	1.16	9.82	9.21	93.0	
M3605-04	PP-1	4	1.16	9.75	9.02	91.5	
M3605-05	PD-1	5	1.14	9.71	8.81	89.5	
M3609-01	EO-01-083121	15	1.11	9.86	9.33	93.9	
M3609-02	EO-01-083121-EPH-2	16	1.12	9.7	9.12	93.2	
M3609-03	EO-03-083121	17	1.13	9.6	9.06	93.6	
M3609-04	EO-03-083121-EPH-2	18	1.16	9.75	9.19	93.5	
M3612-01	M-1	6	1.12	9.83	3.04	22.0	SLUDGE SAMPLE
M3613-01	1	7	1.13	9.51	9.14	95.6	
M3613-02	2	8	1.16	9.62	8.6	87.9	
M3613-03	3	9	1.15	9.96	9.00	89.1	
M3613-04	4	10	1.15	9.6	8.57	87.8	
M3613-05	5	11	1.14	9.97	9.02	89.2	
M3613-06	6	12	1.17	9.83	8.74	87.4	
M3614-01	NB-8-083121	19	1.12	9.67	9.07	93.0	
M3614-02	NB-8-083121-EPH-2	20	1.17	9.81	9.36	94.8	
M3614-02D	NB-8-083121-EPH-2DUP	21	1.19	9.83	9.39	94.9	
M3615-01	TPR-10-COMP	13	1.17	9.8	8.64	86.6	
M3615-02	TPR-10-EPH-2	14	1.19	9.84	8.36	82.9	
M3616-01	CL-01-083121	22	1.12	9.67	9.19	94.4	
M3616-02	CL-01-083121-EPH-2	23	1.14	9.75	9.25	94.2	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$