



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB116096

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	08/31/2021	16:15
2	CAL2	1	Water	NA	NA	21.2	7.00	08/31/2021	16:16
3	CAL3	1	Water	NA	NA	21.3	10.03	08/31/2021	16:19
4	ICV	1	Water	NA	NA	21.5	7.01	08/31/2021	16:20
5	CCV1	1	Water	NA	NA	21.4	2.01	08/31/2021	16:22
6	M3611-01	1	Water	NA	NA	20.1	7.68	08/31/2021	16:24
7	M3611-03	1	Water	NA	NA	22.5	7.00	08/31/2021	16:30
8	M3611-03DUP	1	Water	NA	NA	22.6	7.00	08/31/2021	16:31
9	CCV2	1	Water	NA	NA	21.3	12.03	08/31/2021	16:35

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M3611 WorkList ID : 151912 Department : Wet-Chemistry Date : 08-31-2021 16:07:24

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
08/31/2021	Water	M3611-01	pH	Cool 4 deg C	PSEG03	H12	NWB-1528	08/31/2021	9040C
08/31/2021	Water	M3611-03	pH	Cool 4 deg C	PSEG03	H12	NWB-1601	08/31/2021	9040C

Date/Time 08/31/21 16:10
 Raw Sample Received by: JRC
 Raw Sample Relinquished by: JRC

Date/Time 08/31/21 17:20
 Raw Sample Received by: JRC
 Raw Sample Relinquished by: JRC