


LB116104

245.1

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.002	-0.002	58	0.000	58					
0.05	0.050	0.041	-0.009	212	0.0 %	212	0				-20
0.2	0.200	0.202	0.002	777	0.0 %	777					1
2.5	2.500	2.502	0.002	8888	0.0 %	8888					0
5.0	5.000	4.979	-0.021	17621	0.0 %	17621					0
7.5	7.500	7.579	0.079	26788	0.0 %	26788					1
10.0	10.000	9.950	-0.050	35151	0.0 %	35151					-1

LB116104 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 01 Sep 2021 12:29:30

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	58	-	0.0000	245.1	PPB	01 Sep 2021 12:30:53	S	Std
0.05 - 1	50.05	212	-	0.0500	245.1	PPB	01 Sep 2021 12:33:10	S	Std
0.2 - 1	50.2	777	-	0.2000	245.1	PPB	01 Sep 2021 12:35:27	S	Std
2.5 - 1	52.5	8888	-	2.5000	245.1	PPB	01 Sep 2021 12:37:43	S	Std
5.0 - 1	55	17621	-	5.0000	245.1	PPB	01 Sep 2021 12:40:00	S	Std
7.5 - 1	57.5	26788	-	7.5000	245.1	PPB	01 Sep 2021 12:42:16	S	Std
10.0 - 1	510	35151	-	10.0000	245.1	PPB	01 Sep 2021 12:44:31	S	Std
ICV77 - 1	ICV77	13917	3.9282	-245.1		PPB	01 Sep 2021 12:47:19	U	SMPL
ICB77 - 1	ICB77	-30	-0.0273	-245.1		PPB	01 Sep 2021 12:49:34	U	SMPL
CCV76 - 1	CCV76	17502	4.9449	-245.1		PPB	01 Sep 2021 12:51:48	U	SMPL
CCB76 - 1	CCB76	-35	-0.0287	-245.1		PPB	01 Sep 2021 12:54:04	U	SMPL
CRA - 1	CRA	206	0.0396	-245.1		PPB	01 Sep 2021 12:56:19	U	SMPL
HighStd - 1	HighStd	34689	9.8193	-245.1		PPB	01 Sep 2021 12:58:34	U	SMPL
ChkStd - 1	ChkStd	24646	6.9710	-245.1		PPB	01 Sep 2021 13:00:49	U	SMPL
PB138872BL - 1	PBW	-65	-0.0373	-245.1		PPB	01 Sep 2021 13:03:06	U	SMPL
PB138872BS - 1	LCSW	13995	3.9503	-245.1		PPB	01 Sep 2021 13:05:23	U	SMPL
M3610-01 - 1	YORK-ST-MH-1	41202	11.6665	-245.1		PPB	01 Sep 2021 13:07:39	U	SMPL
M3610-02 - 1	YORK-ST-MH-2	-86	-0.0432	-245.1		PPB	01 Sep 2021 13:09:55	U	SMPL
M3610-02DUP - 1	YORK-ST-MH-2DUP	44	-0.0063	-245.1		PPB	01 Sep 2021 13:12:14	U	SMPL
M3610-02MS - 1	YORK-ST-MH-2MS	14216	4.0130	-245.1		PPB	01 Sep 2021 13:14:31	U	SMPL
M3610-02MSD - 1	YORK-ST-MH-2MSD	14260	4.0255	-245.1		PPB	01 Sep 2021 13:16:46	U	SMPL
CCV77 - 1	CCV77	17948	5.0714	-245.1		PPB	01 Sep 2021 13:19:01	U	SMPL
CCB77 - 1	CCB77	-32	-0.0279	-245.1		PPB	01 Sep 2021 13:21:16	U	SMPL
M3610-02LX5 - 1		29	-0.0106	-245.1		PPB	01 Sep 2021 13:23:32	U	SMPL
M3610-02A - 1		14376	4.0584	-245.1		PPB	01 Sep 2021 13:25:48	U	SMPL
M3610-01DLX2 - 1		18059	5.1029	-245.1		PPB	01 Sep 2021 13:28:03	U	SMPL
CCV78 - 1	CCV78	17491	4.9418	-245.1		PPB	01 Sep 2021 13:30:19	U	SMPL
CCB78 - 1	CCB78	-21	-0.0248	-245.1		PPB	01 Sep 2021 13:32:35	U	SMPL