



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB116113

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	09/01/2021	16:20
2	CAL2	1	Water	NA	NA	21.2	7.00	09/01/2021	16:21
3	CAL3	1	Water	NA	NA	21.2	10.03	09/01/2021	16:22
4	ICV	1	Water	NA	NA	21.3	6.99	09/01/2021	16:24
5	CCV1	1	Water	NA	NA	21.3	2.01	09/01/2021	16:25
6	M3618-01	1	Water	NA	NA	21.5	7.71	09/01/2021	16:30
7	M3618-01DUP	1	Water	NA	NA	21.6	7.72	09/01/2021	16:31
8	M3629-01	1	Water	NA	NA	23.4	6.72	09/01/2021	16:32
9	M3630-01	1	Water	NA	NA	20.1	5.80	09/01/2021	16:35
10	M3630-03	1	Water	NA	NA	20.2	5.67	09/01/2021	16:36
11	M3630-05	1	Water	NA	NA	20.1	4.54	09/01/2021	16:40
12	M3630-07	1	Water	NA	NA	20.2	6.59	09/01/2021	16:42
13	M3630-09	1	Water	NA	NA	20.3	5.21	09/01/2021	16:45
14	M3630-11	1	Water	NA	NA	20.4	6.36	09/01/2021	16:46
15	M3630-13	1	Water	NA	NA	20.4	6.76	09/01/2021	16:50
16	CCV2	1	Water	NA	NA	21.3	12.02	09/01/2021	16:51
17	M3630-14	1	Water	NA	NA	20.4	6.84	09/01/2021	16:52
18	CCV3	1	Water	NA	NA	21.3	2.01	09/01/2021	17:00

B116113

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph m3630 WorkList ID : 151947 Department : Wet-Chemistry Date : 09-01-2021 15:27:04

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/01/2021	Water	M3618-01	pH	Cool 4 deg C	PSEG03	H13	33429	09/01/2021	9040C
09/01/2021	Water	M3629-01	pH	Cool 4 deg C	PSEG03	H13	LPA-02	09/01/2021	9040C
09/01/2021	Water	M3630-01	pH	Cool 4 deg C	PSEG03	H13	33414	09/01/2021	9040C
09/01/2021	Water	M3630-03	pH	Cool 4 deg C	PSEG03	H13	33415	09/01/2021	9040C
09/01/2021	Water	M3630-05	pH	Cool 4 deg C	PSEG03	H13	33416	09/01/2021	9040C
09/01/2021	Water	M3630-07	pH	Cool 4 deg C	PSEG03	H13	33417	09/01/2021	9040C
09/01/2021	Water	M3630-09	pH	Cool 4 deg C	PSEG03	H13	33422	09/01/2021	9040C
09/01/2021	Water	M3630-11	pH	Cool 4 deg C	PSEG03	H13	33423	09/01/2021	9040C
09/01/2021	Water	M3630-13	pH	Cool 4 deg C	PSEG03	H13	33418	09/01/2021	9040C
09/01/2021	Water	M3630-14	pH	Cool 4 deg C	PSEG03	H13	33419	09/01/2021	9040C

Date/Time 09.01.21 18:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 09.01.21 16:13
Raw Sample Received by: [Signature] (WC)
Raw Sample Relinquished by: [Signature] (WC)