

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB116122

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.3	09/02/2021	12:20
2	M3610-01		1	10	100.00	>212.0	>212.0	09/02/2021	12:45
3	M3610-01DUP		1	10	100.00	>212.0	>212.0	09/02/2021	13:05
4	M3610-02		1	12	100.00	>212.0	>212.0	09/02/2021	13:35
5	M3611-01		1	15	100.00	>212.0	>212.0	09/02/2021	13:58
6	M3611-03		1	13	100.00	>212.0	>212.0	09/02/2021	14:20
7	M3618-01		1	11	100.00	>212.0	>212.0	09/02/2021	14:40
8	M3630-01		1	14	100.00	>212.0	>212.0	09/02/2021	15:05
9	M3630-03		1	12	100.00	>212.0	>212.0	09/02/2021	15:25
10	M3630-05		1	14	100.00	>212.0	>212.0	09/02/2021	15:45
11	M3630-07		1	13	100.00	>212.0	>212.0	09/02/2021	16:05
12	M3630-09		1	14	100.00	>212.0	>212.0	09/02/2021	16:32
13	M3630-11		1	13	100.00	>212.0	>212.0	09/02/2021	16:58
14	M3630-15		1	14	100.00	>212.0	>212.0	09/02/2021	17:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 116122

WorkList Name : FP9121 WorkList ID : 151946 Department : Wet-Chemistry Date : 09-01-2021 15:26:21

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/03/2021	Water	M3610-01	Flash Point	Cool 4 deg C	DIST01	D11	YORK-ST-MH-1	08/31/2021	1010B
09/03/2021	Water	M3610-02	Flash Point	Cool 4 deg C	DIST01	D11	YORK-ST-MH-2	08/31/2021	1010B
09/13/2021	Water	M3611-01	Flash Point	Cool 4 deg C	PSEG03	H12	NWB-1528	08/31/2021	1010B
09/13/2021	Water	M3611-03	Flash Point	Cool 4 deg C	PSEG03	H12	NWB-1601	08/31/2021	1010B
09/08/2021	Water	M3618-01	Flash Point	Cool 4 deg C	PSEG03	H13	33429	09/01/2021	1010B
09/08/2021	Water	M3630-01	Flash Point	Cool 4 deg C	PSEG03	H13	33414	09/01/2021	1010B
09/08/2021	Water	M3630-03	Flash Point	Cool 4 deg C	PSEG03	H13	33415	09/01/2021	1010B
09/08/2021	Water	M3630-05	Flash Point	Cool 4 deg C	PSEG03	H13	33416	09/01/2021	1010B
09/08/2021	Water	M3630-07	Flash Point	Cool 4 deg C	PSEG03	H13	33417	09/01/2021	1010B
09/08/2021	Water	M3630-09	Flash Point	Cool 4 deg C	PSEG03	H13	33422	09/01/2021	1010B
09/08/2021	Water	M3630-11	Flash Point	Cool 4 deg C	PSEG03	H13	33423	09/01/2021	1010B
09/08/2021	Water	M3630-15	Flash Point	Cool 4 deg C	PSEG03	H13	418-419	09/01/2021	1010B

Date/Time 09/02/21 12:00
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: JP (WC)

Date/Time 09/02/21 17:25
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: KS (WC)