


LB116123

SFAM01.1

INSTRUMENT ID: CX1



Linear

A= 0.0000e+000

B= 2.7678e-004 slope

C= -8.5703e-003 + intercept

Rho= 0.9999003

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0)
0.00	0.000	0.006	0.006	52	0.000	52					-14
0.05	0.050	0.043	-0.007	188	0.0 %	188					4
0.20	0.200	0.208	0.008	782	0.0 %	782					1
2.50	2.500	2.526	0.026	9159	0.0 %	9159					0
5.00	5.000	5.007	0.007	18122	0.0 %	18122					-2
7.50	7.500	7.386	-0.114	26717	0.0 %	26717					1
10.0	10.000	10.073	0.073	36426	0.0 %	36426					

LB116123
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	52	-		0 SFAM01.1	PPB	9/2/2021 14:20	Std
	0.05 S01	188	-		0.05 SFAM01.1	PPB	9/2/2021 14:23	Std
	0.2 S02	782	-		0.2 SFAM01.1	PPB	9/2/2021 14:25	Std
	2.5 S03	9159	-		2.5 SFAM01.1	PPB	9/2/2021 14:27	Std
	5 S04	18122	-		5 SFAM01.1	PPB	9/2/2021 14:30	Std
	7.5 S05	26717	-		7.5 SFAM01.1	PPB	9/2/2021 14:32	Std
	10 S06	36426	-		10 SFAM01.1	PPB	9/2/2021 14:39	Std
ICV065	ICV065	13833	3.8201	-	SFAM01.1	PPB	9/2/2021 14:42	SMPL
ICB065	ICB065	-23	-0.0149	-	SFAM01.1	PPB	9/2/2021 14:44	SMPL
CCV086	CCV086	17808	4.9202	-	SFAM01.1	PPB	9/2/2021 14:46	SMPL
CCB086	CCB086	-32	-0.0174	-	SFAM01.1	PPB	9/2/2021 14:49	SMPL
PB138902BL	PBW	42	0.0031	-	SFAM01.1	PPB	9/2/2021 14:51	SMPL
M3263-01	MDL-WATER-QT4-2021-07	207	0.0487	-	SFAM01.1	PPB	9/2/2021 14:53	SMPL
CCV087	CCV087	17877	4.9393	-	SFAM01.1	PPB	9/2/2021 14:55	SMPL
CCB087	CCB087	-15	-0.0127	-	SFAM01.1	PPB	9/2/2021 14:58	SMPL