


LB116126

SFAM01.1

INSTRUMENT ID: CY2



Linear

A= 0.0000e+000

B= 2.6309e-004 slope

C= -5.7302e-002 y-intercept

Rho= 0.9999626

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.024	0.024	308	0.000	308					-4
0.05	0.050	0.048	-0.002	402	0.0 %	402					1
0.20	0.200	0.202	0.002	985	0.0 %	985					-2
2.50	2.500	2.441	-0.059	9496	0.0 %	9496					6
5.00	5.000	5.020	0.020	19299	0.0 %	19299					1
7.50	7.500	7.544	0.044	28892	0.0 %	28892					0
10.0	10.000	9.971	-0.029	38118	0.0 %	38118					0

LB116126

INSTRUMENT IOD : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	308	-		0 SFAM01.1	PPB	9/2/2021 15:37	Std
	0.05 S01	402	-	0.05	SFAM01.1	PPB	9/2/2021 15:39	Std
	0.2 S02	985	-	0.2	SFAM01.1	PPB	9/2/2021 15:41	Std
	2.5 S03	9496	-	2.5	SFAM01.1	PPB	9/2/2021 15:44	Std
	5 S04	19299	-	5	SFAM01.1	PPB	9/2/2021 15:46	Std
	7.5 S05	28892	-	7.5	SFAM01.1	PPB	9/2/2021 15:48	Std
	10 S06	38118	-	10	SFAM01.1	PPB	9/2/2021 15:50	Std
ICV066	ICV066	14720	3.8154	-	SFAM01.1	PPB	9/2/2021 15:53	SMPL
ICB066	ICB066	291	0.0193	-	SFAM01.1	PPB	9/2/2021 15:55	SMPL
CCV088	CCV088	19758	5.1408	-	SFAM01.1	PPB	9/2/2021 15:58	SMPL
CCB088	CCB088	286	0.0179	-	SFAM01.1	PPB	9/2/2021 16:00	SMPL
PB138903BL	PBS903	252	0.009	-	SFAM01.1	PPB	9/2/2021 16:02	SMPL
M3263-03	MDL-SOIL-QT4-2021-07	408	0.05	-	SFAM01.1	PPB	9/2/2021 16:04	SMPL
CCV089	CCV089	19790	5.1492	-	SFAM01.1	PPB	9/2/2021 16:07	SMPL
CCB089	CCB089	282	0.0169	-	SFAM01.1	PPB	9/2/2021 16:09	SMPL