

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC

284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : PS

Instrument ID : Konelab

9/2/2021 13:58

Test: ^{DEP} ~~Total~~ CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	94.033	0.0	0.073	
ICB1	0.274	0.0	0.001	
CCV1	237.880	0.0	0.182	
CCB1	0.140	0.0	0.001	
PB138878BL	0.098	0.0	0.001	
M3593-02	0.102	0.0	0.001	
M3593-02DUP	0.466	0.0	0.001	
CCV2	238.225	0.0	0.182	
CCB2	0.131	0.0	0.001	

N	9
Mean	63.483
SD	103.6276
CV%	163.24

Aquakem v. 7.2AQ1

Results from time period:

Thu Sep 02 12:01:45 2021

Thu Sep 02 13:51:57 2021

Sample Id	Sam/Ctr/c/	Test short nam	Test type	Result	Result unit	Result date and time
0.OPPBCN	A	Reactive CN	P	0.291	µg/l	9/2/2021 13:14:14
5.OPPBCN	A	Reactive CN	P	4.8084	µg/l	9/2/2021 13:14:15
10PPBCN	A	Reactive CN	P	9.0089	µg/l	9/2/2021 13:14:16
50PPBCN	A	Reactive CN	P	50.3232	µg/l	9/2/2021 13:14:17
100PPBCN	A	Reactive CN	P	99.8694	µg/l	9/2/2021 13:14:18
250PPBCN	A	Reactive CN	P	251.3672	µg/l	9/2/2021 13:14:19
500PPBCN	A	Reactive CN	P	499.332	µg/l	9/2/2021 13:14:20
ICV1	S	Reactive CN	P	94.0327	µg/l	9/2/2021 13:40:59
ICB1	S	Reactive CN	P	0.2738	µg/l	9/2/2021 13:41:00
CCV1	S	Reactive CN	P	237.8797	µg/l	9/2/2021 13:41:03
CCB1	S	Reactive CN	P	0.1395	µg/l	9/2/2021 13:41:04
PB138878BL	S	Reactive CN	P	0.098	µg/l	9/2/2021 13:41:06
M3593-02	S	Reactive CN	P	0.1021	µg/l	9/2/2021 13:48:33
M3593-02DUP	S	Reactive CN	P	0.4665	µg/l	9/2/2021 13:48:34
CCV2	S	Reactive CN	P	238.2255	µg/l	9/2/2021 13:48:37
CCB2	S	Reactive CN	P	0.1314	µg/l	9/2/2021 13:48:39

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : AA

Instrument ID : Konelab

9/2/2021 13:16


Test Total CN

Accepted 9/2/2021 13:16

Factor 1313
Bias 0.001

Coeff. of det. 0.999983

Errors



	Calibrator	Response	Calc. con.	Conc.	Rel. Errors
1	0.0PPBCN	0.001	0.2910	0.0000	-3.8
2	5.0PPBCN	0.005	4.8084	5.0000	-9.9
3	10PPBCN	0.008	9.0089	10.0000	0.6
4	50PPBCN	0.039	50.3232	50.0000	0.1
5	100PPBCN	0.077	99.8694	100.0000	0.5
6	250PPBCN	0.192	251.3672	250.0000	-0.1
7	500PPBCN	0.381	499.3320	500.0000	

AB
9/2/21