

TOTAL SOLIDS - SM2540B

SUPERVISOR: apatel

ANALYST: jignesh

Date: 09/03/2021

OVEN IN TEMP1: 104 °C

OVEN OUT TEMP1: 103 °C

DATE & TIME IN1: 09/02/2021 17:10

DATE & TIME OUT1: 09/03/2021 08:00

OVEN IN TEMP2: 103 °C

OVEN OUT TEMP2: 104 °C

DATE & TIME IN2: 09/03/2021 08:30

DATE & TIME OUT2: 09/03/2021 10:00

Run Number: LB116135

BalanceID: Wc Sc-6

OvenID: Wc Oven-1

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB116135BL	LB116135BL	85.6371	85.6371	100	85.6372	85.6372	85.6372	0.0001	1
2	M3644-02	191-0902-W	75.5516	75.5516	100	75.6161	75.6161	75.6161	0.0645	645
3	M3644-02DUP	191-0902-WDUP	85.0833	85.0833	100	85.1456	85.1456	85.1456	0.0623	623
4	M3644-04	191-0902-DUP	68.8278	68.8278	100	68.8917	68.8917	68.8917	0.0639	639

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result mg/L} = ((C - B) / A) * 1000 * 1000$$

HP 116135

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS M3640 WorkList ID : 151975 Department : Wet-Chemistry Date : 09-02-2021 16:42:44

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/09/2021	Water	M3644-02	TS	Cool 4 deg C	ENTE01	D11	191-0902-W	09/02/2021	SM2540 B
09/09/2021	Water	M3644-04	TS	Cool 4 deg C	ENTE01	D11	191-0902-DUP	09/02/2021	SM2540 B

Date/Time 09/02/21 16:50
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: JP (WC)

Date/Time 09/02/21 18:00
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: JP (WC)