



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB116136

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	09/02/2021	16:20
2	CAL2	1	Water	NA	NA	21.2	7.00	09/02/2021	16:21
3	CAL3	1	Water	NA	NA	21.2	10.03	09/02/2021	16:22
4	ICV	1	Water	NA	NA	21.2	7.01	09/02/2021	16:24
5	CCV1	1	Water	NA	NA	21.3	2.01	09/02/2021	16:25
6	M3646-03	1	Water	NA	NA	22.2	6.78	09/02/2021	16:26
7	M3646-04	1	Water	NA	NA	22.4	7.02	09/02/2021	16:30
8	M3646-05	1	Water	NA	NA	22.2	7.36	09/02/2021	16:32
9	M3646-06	1	Water	NA	NA	21.2	6.47	09/02/2021	16:34
10	M3646-07	1	Water	NA	NA	21.4	5.92	09/02/2021	16:35
11	M3646-08	1	Water	NA	NA	21.5	6.40	09/02/2021	16:36
12	M3646-09	1	Water	NA	NA	21.8	6.82	09/02/2021	16:39
13	M3646-10	1	Water	NA	NA	21.0	6.78	09/02/2021	16:40
14	M3646-13	1	Water	NA	NA	20.9	6.95	09/02/2021	16:42
15	M3646-13DUP	1	Water	NA	NA	20.8	6.96	09/02/2021	16:43
16	CCV2	1	Water	NA	NA	21.4	12.02	09/02/2021	16:45

28116136

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M3646 WorkList ID : 151974 Department : Wet-Chemistry Date : 09-02-2021 16:08:05

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/02/2021	Water	M3646-03	pH	Cool 4 deg C	PSEG03	H21	2106	09/02/2021	9040C
09/02/2021	Water	M3646-04	pH	Cool 4 deg C	PSEG03	H21	2107	09/02/2021	9040C
09/02/2021	Water	M3646-05	pH	Cool 4 deg C	PSEG03	H21	2124	09/02/2021	9040C
09/02/2021	Water	M3646-06	pH	Cool 4 deg C	PSEG03	H21	2099	09/02/2021	9040C
09/02/2021	Water	M3646-07	pH	Cool 4 deg C	PSEG03	H21	2100	09/02/2021	9040C
09/02/2021	Water	M3646-08	pH	Cool 4 deg C	PSEG03	H21	2101	09/02/2021	9040C
09/02/2021	Water	M3646-09	pH	Cool 4 deg C	PSEG03	H21	2093	09/02/2021	9040C
09/02/2021	Water	M3646-10	pH	Cool 4 deg C	PSEG03	H21	2094	09/02/2021	9040C
09/02/2021	Water	M3646-13	pH	Cool 4 deg C	PSEG03	H21	1992	09/02/2021	9040C

Date/Time 09/02/21 16:10
Raw Sample Received by: JR (WC)
Raw Sample Relinquished by: JD CSM

Date/Time 09/02/21 17:30
Raw Sample Received by: JD CSM
Raw Sample Relinquished by: JR (WC)