


LB116146

SFAM01.1

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.8823e-004 slope

C= 1.0143e-003 y-intercept

Rho= 0.9998356

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.023	0.023	76	0.000	76					14
0.05	0.050	0.057	0.007	194	0.0 %	194					8
0.20	0.200	0.215	0.015	742	0.0 %	742					0
2.50	2.500	2.510	0.010	8704	0.0 %	8704					-1
5.00	5.000	4.970	-0.030	17239	0.0 %	17239					-2
7.50	7.500	7.366	-0.134	25553	0.0 %	25553					1
10.0	10.000	10.110	0.110	35071	0.0 %	35071					

LB116146
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	76	-		0 SFAM01.1	PPB	9/3/2021 12:45	Std
	0.05 S01	194	-	0.05	SFAM01.1	PPB	9/3/2021 12:48	Std
	0.2 S02	742	-	0.2	SFAM01.1	PPB	9/3/2021 12:50	Std
	2.5 S03	8704	-	2.5	SFAM01.1	PPB	9/3/2021 12:52	Std
	5 S04	17239	-	5	SFAM01.1	PPB	9/3/2021 12:54	Std
	7.5 S05	25553	-	7.5	SFAM01.1	PPB	9/3/2021 12:57	Std
	10 S06	35071	-	10	SFAM01.1	PPB	9/3/2021 12:59	Std
ICV067	ICV067	13911	4.0106	-	SFAM01.1	PPB	9/3/2021 13:02	SMPL
ICB067	ICB067	-4	-0.0001	-	SFAM01.1	PPB	9/3/2021 13:04	SMPL
CCV094	CCV094	16862	4.8611	-	SFAM01.1	PPB	9/3/2021 13:06	SMPL
CCB094	CCB094	-13	-0.0027	-	SFAM01.1	PPB	9/3/2021 13:09	SMPL
PB138933BL	PBW933	40	0.0125	-	SFAM01.1	PPB	9/3/2021 13:11	SMPL
M3263-02	MDL-WATER-QT4-2021-08	218	0.0638	-	SFAM01.1	PPB	9/3/2021 13:13	SMPL
CCV095	CCV095	17288	4.9839	-	SFAM01.1	PPB	9/3/2021 13:15	SMPL
CCB095	CCB095	-6	-0.0007	-	SFAM01.1	PPB	9/3/2021 13:18	SMPL