


LB116151

SFAM01.1

INSTRUMENT ID: CV2



Linear

A= 0.0000e+000

B= 2.6694e-004 slope

C= -6.0970e-002 intercept

Rho= 0.9998801

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	9/01)
0.00	0.000	-0.003	-0.003	216	0.000	216					-16
0.05	0.050	0.042	-0.008	387	0.0 %	387					-2
0.20	0.200	0.196	-0.004	963	0.0 %	963					-2
2.50	2.500	2.455	-0.045	9425	0.0 %	9425					1
5.00	5.000	5.052	0.052	19154	0.0 %	19154					1
7.50	7.500	7.600	0.100	28698	0.0 %	28698					-1
10.0	10.000	9.908	-0.092	37347	0.0 %	37347					

LB116151

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	216	-		0 SFAM01.1	PPB	9/3/2021 14:27	Std
	0.05 S01	387	-	0.05	SFAM01.1	PPB	9/3/2021 14:29	Std
	0.2 S02	963	-	0.2	SFAM01.1	PPB	9/3/2021 14:32	Std
	2.5 S03	9425	-	2.5	SFAM01.1	PPB	9/3/2021 14:34	Std
	5 S04	19154	-	5	SFAM01.1	PPB	9/3/2021 14:36	Std
	7.5 S05	28698	-	7.5	SFAM01.1	PPB	9/3/2021 14:39	Std
	10 S06	37347	-	10	SFAM01.1	PPB	9/3/2021 14:41	Std
ICV068	ICV068	14749	3.8761 -		SFAM01.1	PPB	9/3/2021 14:45	SMPL
ICB068	ICB068	183	-0.0121 -		SFAM01.1	PPB	9/3/2021 14:48	SMPL
CCV096	CCV096	19041	5.0218 -		SFAM01.1	PPB	9/3/2021 14:50	SMPL
CCB096	CCB096	182	-0.0124 -		SFAM01.1	PPB	9/3/2021 14:52	SMPL
PB138934BL	PBS934	213	-0.0041 -		SFAM01.1	PPB	9/3/2021 14:54	SMPL
M3263-04	MDL-SOIL-QT4-2021-08	373	0.0386 -		SFAM01.1	PPB	9/3/2021 14:57	SMPL
CCV097	CCV097	19259	5.08 -		SFAM01.1	PPB	9/3/2021 15:02	SMPL
CCB097	CCB097	176	-0.014 -		SFAM01.1	PPB	9/3/2021 15:04	SMPL