



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB116153

Slope : 98.1

BalanceID: Wc Sc-06

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	09/03/2021	14:45
2	CAL2	1	Water	NA	NA	21.2	7.00	09/03/2021	14:46
3	CAL3	1	Water	NA	NA	21.2	10.03	09/03/2021	14:47
4	ICV	1	Water	NA	NA	21.2	7.02	09/03/2021	14:49
5	CCV1	1	Water	NA	NA	21.2	2.01	09/03/2021	14:50
6	M3655-04	1	Solid	20.03	20	22.4	11.12	09/03/2021	15:00
7	M3655-04DUP	1	Solid	20.04	20	22.5	11.12	09/03/2021	15:01
8	CCV2	1	Water	NA	NA	21.2	12.02	09/03/2021	15:03

WB 116153

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity m3655 WorkList ID : 152009 Department : Wet-Chemistry Date : 09-03-2021 14:39:00

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/03/2021	Solid	M3655-04	✓ Corrosivity	Cool 4 deg C	PSEG03	H22	CON-GROWS	09/03/2021	9045D

Date/Time 09/03/21 14:40
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: CP SM

Date/Time 09/03/21 16:00
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: JP (WC)