



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatel

ANALYST: jignesh

Date: 09/07/2021

OVEN IN TEMP1: 104 °C

OVEN OUT TEMP1: 103 °C

DATE & TIME IN1: 09/03/2021 16:00

DATE & TIME OUT1: 09/07/2021 08:00

OVEN IN TEMP2: 104 °C

OVEN OUT TEMP2: 104 °C

DATE & TIME IN2: 09/07/2021 08:30

DATE & TIME OUT2: 09/07/2021 10:00

Run Number: LB116154

BalanceID: Wc Sc-06

OvenID: Wc Oven-01

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Dish + Sample Weight (g)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result %
1	LB116154BL	LB116154BL	84.5371	84.5371	95.1473	95.1474	95.1474	95.1474	10.6103	0
2	M3655-03	CON-GROWS	59.3967	59.3967	81.7742	80.8917	80.8917	80.8917	21.4950	96.1
3	M3655-03DUP	CON-GROWSDUP	53.5558	53.5558	76.901	75.7296	75.7296	75.7296	22.1738	95

A = Final Empty Dish Weight (g)

B = Dish + Sample Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

Result % = (C - A) * 100 / (B - A)

2B 116154

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ts m3655 s WorkList ID : 152010 Department : Wet-Chemistry Date : 09-03-2021 14:40:02

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/09/2021	Solid	M3655-03	✓ TS	Cool 4 deg C	PSEG03	H22	CON-GROWS	09/03/2021	SM2540 B

Date/Time 09/03/21 14:45
Raw Sample Received by: JP (WC)
Raw Sample Relinquished by: JP SM

Date/Time 09/03/21 16:10
Raw Sample Received by: CP SM
Raw Sample Relinquished by: JP (WC)