



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB116166

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	09/07/2021	10:45
2	M3629-01		1	10	100.00	>212.0	>212.0	09/07/2021	11:10
3	M3629-01DUP		1	12	100.00	>212.0	>212.0	09/07/2021	11:40
4	M3644-02		1	10	100.00	>212.0	>212.0	09/07/2021	12:05
5	M3644-04		1	12	100.00	>212.0	>212.0	09/07/2021	12:35
6	M3646-11		1	15	100.00	>212.0	>212.0	09/07/2021	12:58

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 116166

WorkList Name : fp9721

WorkList ID : 152024

Department : Wet-Chemistry

Date : 09-07-2021 08:47:20

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/06/2021	Water	M3629-01	C Flash Point	Cool 4 deg C	PSEG03	H13	LPA-02	09/01/2021	1010B
09/16/2021	Water	M3644-02	P Flash Point	Cool 4 deg C	ENTE01	D11	191-0902-W	09/02/2021	1010B
09/16/2021	Water	M3644-04	P Flash Point	Cool 4 deg C	ENTE01	D11	191-0902-DUP	09/02/2021	1010B
09/09/2021	Water	M3646-11	C Flash Point	Cool 4 deg C	PSEG03	H21	2105	09/02/2021	1010B

Date/Time 09/07/21 08:55
Raw Sample Received by: KS(WC)
Raw Sample Relinquished by: JS(WC)

Date/Time 09/07/21 13:00
Raw Sample Received by: JS(WC)
Raw Sample Relinquished by: JS(WC)