

LB116170


7470A

INSTRUMENT ID: CVL

Linear

 μ Abs.:

34727



A= 0.0000e+000

B= 2.8665e-004

C= -1.4885e-003

Rho= 0.9996170

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.063	0.063	226	0.000	226					0
0.2	0.200	0.200	0.000	703	0.0 %	703					-1
2.5	2.500	2.475	-0.025	8641	0.0 %	8641					-3
5.0	5.000	4.837	-0.163	16880	0.0 %	16880					2
7.5	7.500	7.671	0.171	26766	0.0 %	26766					0
10.0	10.000	9.953	-0.047	34727	0.0 %	34727					

LB116170 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 07 Sep 2021 14:52:13

Sample ID	Extended ID	µ Abs	Conc	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	226	-	0.00007470A	PPB	07 Sep 2021 14:55:21	S	Std
0.2 - 1	50.2	703	-	0.20007470A	PPB	07 Sep 2021 14:57:38	S	Std
2.5 - 1	52.5	8641	-	2.50007470A	PPB	07 Sep 2021 14:59:55	S	Std
5.0 - 1	55	16880	-	5.00007470A	PPB	07 Sep 2021 15:02:12	S	Std
7.5 - 1	57.5	26766	-	7.50007470A	PPB	07 Sep 2021 15:04:28	S	Std
10.0 - 1	516	34727	-	10.00007470A	PPB	07 Sep 2021 15:09:32	S	Std
ICV84 - 1	ICV84	14638	4.1945	-7470A	PPB	07 Sep 2021 15:12:17	U	SMPL
ICB84 - 1	ICB84	127	0.0349	-7470A	PPB	07 Sep 2021 15:14:31	U	SMPL
CCV07 - 1	CCV07	16765	4.8042	-7470A	PPB	07 Sep 2021 15:16:46	U	SMPL
CCB07 - 1	CCB07	139	0.0384	-7470A	PPB	07 Sep 2021 15:19:02	U	SMPL
CRA - 1	CRA	687	0.1954	-7470A	PPB	07 Sep 2021 15:21:18	U	SMPL
HighStd - 1	HighStd	35253	10.1038	-7470A	PPB	07 Sep 2021 15:23:33	U	SMPL
ChkStd - 1	ChkStd	25944	7.4354	-7470A	PPB	07 Sep 2021 15:25:50	U	SMPL
PB138961BL - 1	PBW	76	0.0203	-7470A	PPB	07 Sep 2021 15:28:07	U	SMPL
PB138961BS - 1	LCSW	15053	4.3135	-7470A	PPB	07 Sep 2021 15:30:23	U	SMPL
M3655-04 - 1	CON-GROWS	102	0.0277	-7470A	PPB	07 Sep 2021 15:32:40	U	SMPL
M3655-04DUP - 1	CON-GROWSDUP	148	0.0409	-7470A	PPB	07 Sep 2021 15:34:57	U	SMPL
M3655-04MS - 1	CON-GROWSMS	12776	3.6608	-7470A	PPB	07 Sep 2021 15:37:14	U	SMPL
M3655-04MSD - 1	CON-GROWSMSD	14517	4.1598	-7470A	PPB	07 Sep 2021 15:39:31	U	SMPL
M3659-01 - 1	TREMONT-AVE-STATION-DRUM-10-PPE	115	0.0315	-7470A	PPB	07 Sep 2021 15:41:46	U	SMPL
CCV08 - 1	CCV08	16660	4.7741	-7470A	PPB	07 Sep 2021 15:44:01	U	SMPL
CCB08 - 1	CCB08	152	0.0421	-7470A	PPB	07 Sep 2021 15:46:16	U	SMPL
M3659-02 - 1	TREMONT-AVE-STATION-DRUM-11-POLY	195	0.0544	-7470A	PPB	07 Sep 2021 15:48:31	U	SMPL
PB138961TB - 1	PB138938TB	152	0.0421	-7470A	PPB	07 Sep 2021 15:50:46	U	SMPL
M3655-04LX5 - 1		15	0.0028	-7470A	PPB	07 Sep 2021 15:53:02	U	SMPL
M3655-04A - 1		13271	3.8027	-7470A	PPB	07 Sep 2021 15:55:18	U	SMPL
CCV09 - 1	CCV09	16139	4.6248	-7470A	PPB	07 Sep 2021 16:00:09	U	SMPL
CCB09 - 1	CCB09	124	0.0341	-7470A	PPB	07 Sep 2021 16:02:24	U	SMPL