



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB116171

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	09/07/2021	13:15
2	CAL2	1	Water	NA	NA	21.3	7.00	09/07/2021	13:16
3	CAL3	1	Water	NA	NA	21.3	10.03	09/07/2021	13:18
4	ICV	1	Water	NA	NA	21.3	6.99	09/07/2021	13:20
5	CCV1	1	Water	NA	NA	21.4	2.01	09/07/2021	13:22
6	M3666-05	1	Water	NA	NA	21.0	9.60	09/07/2021	13:25
7	M3666-05DUP	1	Water	NA	NA	21.1	9.61	09/07/2021	13:26
8	CCV2	1	Water	NA	NA	21.4	12.03	09/07/2021	13:30

1516171

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph m3666 WorkList ID : 152040 Department : Wet-Chemistry Date : 09-07-2021 13:04:58

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/07/2021	Water	M3666-05	pH	Cool 4 deg C	MAJO01	D11	DRAIN-WATER-TANK-1	09/07/2021	9040C

Date/Time 09/07/21 13:10
Raw Sample Received by: JR WOC
Raw Sample Relinquished by: CP SM

Date/Time 09/07/21 16:20
Raw Sample Received by: CP SM
Raw Sample Relinquished by: JR WOC