



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB116177

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.9	09/07/2021	09:20
2	M3646-01		1	8	56.00	132.8	133.1	09/07/2021	09:50
3	M3646-01DUP		1	8	57.00	134.6	134.9	09/07/2021	10:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB116177

WorkList Name : fp9721-1 WorkList ID : 152025 Department : Wet-Chemistry Date : 09-07-2021 08:48:17

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/09/2021	Solid	M3646-01	Flash Point	Cool 4 deg C	PSEG03	H21	2104	09/02/2021	1010A

Date/Time 09/07/21 08:55
 Raw Sample Received by: KS (WOC)
 Raw Sample Relinquished by: JP (WOC)

Date/Time 09/07/21 10:15
 Raw Sample Received by: JP (WOC)
 Raw Sample Relinquished by: KS (WOC)