



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB116183

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	09/08/2021	11:35
2	CAL2	1	Water	NA	NA	20.3	7.00	09/08/2021	11:36
3	CAL3	1	Water	NA	NA	20.3	10.03	09/08/2021	11:37
4	ICV	1	Water	NA	NA	20.4	6.99	09/08/2021	11:40
5	CCV1	1	Water	NA	NA	20.4	2.01	09/08/2021	11:43
6	M3678-01	1	Water	NA	NA	20.1	4.90	09/08/2021	11:45
7	M3678-01DUP	1	Water	NA	NA	20.2	4.91	09/08/2021	11:46
8	CCV2	1	Water	NA	NA	20.4	12.02	09/08/2021	11:49

23116183

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M3678 WorkList ID : 152053 Department : Wet-Chemistry Date : 09-08-2021 11:25:19

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/07/2021	Water	M3678-01	pH	Cool 4 deg C	STER12	H31	CELL-POOL-WATER-3RI	09/07/2021	9040C

Date/Time 09/08/21 11:30
Raw Sample Received by: JP (WC)
Raw Sample Relinquished by: JP SM

Date/Time 09/08/21 12:20
Raw Sample Received by: CR SM
Raw Sample Relinquished by: JP (WC)