



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB116187

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	09/08/2021	13:40
2	CAL2	1	Water	NA	NA	21.2	7.00	09/08/2021	13:41
3	CAL3	1	Water	NA	NA	21.2	10.03	09/08/2021	13:42
4	ICV	1	Water	NA	NA	21.3	6.99	09/08/2021	13:45
5	CCV1	1	Water	NA	NA	21.3	2.01	09/08/2021	13:47
6	M3679-05	1	Water	NA	NA	20.4	6.37	09/08/2021	13:50
7	M3679-05DUP	1	Water	NA	NA	20.5	6.38	09/08/2021	13:51
8	CCV2	1	Water	NA	NA	21.2	12.02	09/08/2021	13:52

VB 116187

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M3679

WorkList ID : 152063

Department : Wet-Chemistry

Date : 09-08-2021 13:33:55

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/08/2021	Water	M3679-05 ✓	pH	Cool 4 deg C	MAJO01	D11	DRAIN-WATER-TANK-1	09/08/2021	9040C

Date/Time 09/08/21 13:35
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: CP SM

Date/Time 09/08/21 14:35
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: JP (WC)