

LB116191


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

41658



A= 0.0000e+000

B= 2.3920e-004 slope

C= 1.0538e-003 y-intercept

Rho= 0.9998650

Accept = Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.014	0.014	55	0.000	55					=
0.05	0.050					387					-15
0.20	0.200	0.170	-0.030	707	0.0 %	707					-2
2.50	2.500	2.446	-0.054	10220	0.0 %	10220					3
5.00	5.000	5.128	0.128	21432	0.0 %	21432					0
7.50	7.500	7.477	-0.023	31252	0.0 %	31252					0
10.0	10.000	9.966	-0.034	41658	0.0 %	41658					0

LB116191

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	55	-		0 SFAM01.1	PPB	9/8/2021 15:16	Std
	0.2 S01	707	-		0.2 SFAM01.1	PPB	9/8/2021 15:18	Std
	2.5 S02	10220	-		2.5 SFAM01.1	PPB	9/8/2021 15:21	Std
	5 S03	21432	-		5 SFAM01.1	PPB	9/8/2021 15:23	Std
	7.5 S04	31252	-		7.5 SFAM01.1	PPB	9/8/2021 15:25	Std
	10 S05	41658	-		10 SFAM01.1	PPB	9/8/2021 15:28	Std
ICV069	ICV069	15824	3.7862 -		SFAM01.1	PPB	9/8/2021 15:30	SMPL
ICB069	ICB069	-3	0.0003 -		SFAM01.1	PPB	9/8/2021 15:33	SMPL
CCV013	CCV013	21266	5.0879 -		SFAM01.1	PPB	9/8/2021 15:35	SMPL
CCB013	CCB013	-21	-0.004 -		SFAM01.1	PPB	9/8/2021 15:37	SMPL
PB138977BL	PBW977	30	0.0082 -		SFAM01.1	PPB	9/8/2021 15:39	SMPL
M3619-01	MC0AA1	81	0.0204 -		SFAM01.1	PPB	9/8/2021 15:42	SMPL
M3619-02	MC0AA1D	60	0.0154 -		SFAM01.1	PPB	9/8/2021 15:44	SMPL
M3619-03	MC0AA1S	3538	0.8473 -		SFAM01.1	PPB	9/8/2021 15:46	SMPL
M3619-04	MC0AA3	90	0.0226 -		SFAM01.1	PPB	9/8/2021 15:48	SMPL
CCV014	CCV014	20979	5.0193 -		SFAM01.1	PPB	9/8/2021 15:51	SMPL
CCB014	CCB014	-18	-0.0033 -		SFAM01.1	PPB	9/8/2021 15:53	SMPL