



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB116192

Slope : 98.3

BalanceID: Wc Sc-06

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	09/07/2021	15:30
2	CAL2	1	Water	NA	NA	21.1	7.00	09/07/2021	15:31
3	CAL3	1	Water	NA	NA	21.1	10.03	09/07/2021	15:32
4	ICV	1	Water	NA	NA	21.2	6.98	09/07/2021	15:35
5	CCV1	1	Water	NA	NA	21.1	2.01	09/07/2021	15:36
6	M3669-01	1	Solid	20.03	20	22.2	4.55	09/07/2021	15:40
7	M3670-01	1	Solid	20.02	20	22.0	8.70	09/07/2021	15:42
8	M3671-01	1	Solid	20.03	20	22.2	7.40	09/07/2021	15:45
9	M3671-01DUP	1	Solid	20.04	20	22.3	7.41	09/07/2021	15:46
10	CCV2	1	Water	NA	NA	21.1	12.03	09/07/2021	15:48

WORKLIST(Hardcopy Internal Chain)

13116142

WorkList Name : PH M3671 WorkList ID : 152068 Department : Wet-Chemistry Date : 09-07-2021 15:20:54

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/07/2021	Solid	M3669-01	pH	Cool 4 deg C	FREE02	H23	UTILITY-POLES-BRIGE-I	09/07/2021	9045D
09/07/2021	Solid	M3670-01	pH	Cool 4 deg C	FREE02	H23	SAND-BLAST-MEDIA	09/07/2021	9045D
09/07/2021	Solid	M3671-01	pH	Cool 4 deg C	FREE02	H23	FILTER-MEDIA-FILTERS	09/07/2021	9045D

Date/Time 09/07/21 15:22
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: CP SM

Date/Time 09/07/21 16:30
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: JP (WC)