



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB116198

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	09/08/2021	17:20
2	CAL2	1	Water	NA	NA	20.6	7.00	09/08/2021	17:21
3	CAL3	1	Water	NA	NA	20.6	10.03	09/08/2021	17:22
4	ICV	1	Water	NA	NA	20.7	6.99	09/08/2021	17:25
5	CCV1	1	Water	NA	NA	20.6	2.01	09/08/2021	17:26
6	M3681-04	1	Water	NA	NA	20.4	7.60	09/08/2021	17:29
7	M3681-04DUP	1	Water	NA	NA	20.5	7.61	09/08/2021	17:30
8	CCV2	1	Water	NA	NA	20.7	12.02	09/08/2021	17:31

WORKLIST(Hardcopy Internal Chain)

VP 116198

WorkList Name : PH M3681 W WorkList ID : 152076 Department : Wet-Chemistry Date : 09-08-2021 17:14:11

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/08/2021	Water	M3681-04	pH	Cool 4 deg C	PSEG03	H33	WATER-PIT	09/08/2021	9040C

Date/Time 09/08/21 17:15
 Raw Sample Received by: JP/WC
 Raw Sample Relinquished by: KS (WC)

Date/Time 09/08/21 18:20
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: JP (WC)