



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatelANALYST: jigneshDate: 09/09/2021OVEN IN TEMP1: 104 °COVEN OUT TEMP1: 103 °CDATE & TIME IN1: 09/08/2021 17:30 DATE & TIME OUT1: 09/09/2021 08:00OVEN IN TEMP2: 104 °COVEN OUT TEMP2: 104 °CDATE & TIME IN2: 09/09/2021 08:30 DATE & TIME OUT2: 09/09/2021 10:00Run Number: LB116199BalanceID: Wc Sc-05OvenID: Wc Oven-01

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB116199BL	LB116199BL	145.3087	145.3087	100	145.3088	145.3088	145.3088	0.0001	1
2	M3681-04	WATER-PIT	152.9673	152.9673	100	153.0120	153.0120	153.0120	0.0447	447
3	M3681-04DUP	WATER-PITDUP	157.6167	157.6167	100	157.6618	157.6618	157.6618	0.0451	451

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result mg/L} = ((C - B) / A) * 1000 * 1000$$

VB 116199

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS M3681 W WorkList ID : 152077 Department : Wet-Chemistry Date : 09-08-2021 17:14:54

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/13/2021	Water	M3681-04	TS	Cool 4 deg C	PSEG03	H33	WATER-PIT	09/08/2021	SM2540 B

Date/Time 09/08/21 17:15
Raw Sample Received by: JP (WC)
Raw Sample Relinquished by: KS (WC)

Date/Time 09/08/21 18:20
Raw Sample Received by: KS (WC)
Raw Sample Relinquished by: JP (WC)