


LB116209

7471B

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.0828e-004

C= -2.6857e-002

Rho= 0.9996848

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%)
0.0	0.000	-0.021	-0.021	29	0.000	29					
0.2	0.200	0.186	-0.014	1023	0.0 %	1023					
2.5	2.500	2.512	0.012	12188	0.0 %	12188					
5.0	5.000	5.145	0.145	24831	0.0 %	24831					
7.5	7.500	7.333	-0.167	35339	0.0 %	35339					
10.0	10.000	10.045	0.045	48357	0.0 %	48357					

%)

-7

0

3

-2

0

LB116209 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 09 Sep 2021 12:01:21

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	29	-	0.0000	7471B	PPB	09 Sep 2021 12:02:29	S	Std
0.2 - 1	50-2	1023	-	0.2000	7471B	PPB	09 Sep 2021 12:04:46	S	Std
2.5 - 1	52-5	12188	-	2.5000	7471B	PPB	09 Sep 2021 12:07:03	S	Std
5.0 - 1	55	24831	-	5.0000	7471B	PPB	09 Sep 2021 12:09:19	S	Std
7.5 - 1	57.5	35339	-	7.5000	7471B	PPB	09 Sep 2021 12:11:35	S	Std
10.0 - 1	510	48357	-	10.0000	7471B	PPB	09 Sep 2021 12:16:27	S	Std
ICV86 - 1	ICV86	20136	4.1670	-	7471B	PPB	09 Sep 2021 12:19:04	U	SMPL
ICB86 - 1	ICB86	-35	-0.0341	-	7471B	PPB	09 Sep 2021 12:21:19	U	SMPL
CCV15 - 1	CCV15	24833	5.1453	-	7471B	PPB	09 Sep 2021 12:23:33	U	SMPL
CCB15 - 1	CCB15	-54	-0.0381	-	7471B	PPB	09 Sep 2021 12:25:48	U	SMPL
CRA - 1	CRA	1000	0.1814	-	7471B	PPB	09 Sep 2021 12:28:04	U	SMPL
HighStd - 1	HighStd	47956	9.9612	-	7471B	PPB	09 Sep 2021 12:30:19	U	SMPL
ChkStd - 1	ChkStd	33206	6.8892	-	7471B	PPB	09 Sep 2021 12:32:34	U	SMPL
PB138999BL - 1	PBS	-51	-0.0375	-	7471B	PPB	09 Sep 2021 12:34:51	U	SMPL
PB138999BS - 1	LCSS	20101	4.1597	-	7471B	PPB	09 Sep 2021 12:39:41	U	SMPL
M3668-01 - 1	SP-1	1095	0.2012	-	7471B	PPB	09 Sep 2021 12:41:56	U	SMPL
M3680-01 - 1	FK-SF-02-006-A	3722	0.7483	-	7471B	PPB	09 Sep 2021 12:44:13	U	SMPL
M3680-03 - 1	FK-SF-02-006-B	5162	1.0483	-	7471B	PPB	09 Sep 2021 12:46:30	U	SMPL
M3681-01 - 1	BS-STOCK	1393	0.2633	-	7471B	PPB	09 Sep 2021 12:48:46	U	SMPL
M3681-01DUP - 1	BS-STOCKDUP	1326	0.2493	-	7471B	PPB	09 Sep 2021 12:53:38	U	SMPL
CCV16 - 1	CCV16	24879	5.1548	-	7471B	PPB	09 Sep 2021 12:55:53	U	SMPL
CCB16 - 1	CCB16	-32	-0.0335	-	7471B	PPB	09 Sep 2021 12:58:08	U	SMPL
M3681-01MS - 1	BS-STOCKMS	20563	4.2559	-	7471B	PPB	09 Sep 2021 13:00:23	U	SMPL
M3681-01MSD - 1	BS-STOCKMSD	23021	4.7679	-	7471B	PPB	09 Sep 2021 13:02:38	U	SMPL
M3681-01LX5 - 1		231	0.0213	-	7471B	PPB	09 Sep 2021 13:04:54	U	SMPL
M3681-01A - 1		21416	4.4336	-	7471B	PPB	09 Sep 2021 13:07:09	U	SMPL
CCV17 - 1	CCV17	24768	5.1317	-	7471B	PPB	09 Sep 2021 13:09:25	U	SMPL
CCB17 - 1	CCB17	-36	-0.0344	-	7471B	PPB	09 Sep 2021 13:11:41	U	SMPL