



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB116215

Slope : 98.6

BalanceID: Wc Sc-06

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	09/09/2021	13:40
2	CAL2	1	Water	NA	NA	20.3	7.00	09/09/2021	13:41
3	CAL3	1	Water	NA	NA	20.3	10.02	09/09/2021	13:42
4	ICV	1	Water	NA	NA	20.3	7.02	09/09/2021	13:45
5	CCV1	1	Water	NA	NA	20.4	2.01	09/09/2021	13:46
6	M3686-01	1	Solid	20.03	20	21.7	12.15	09/09/2021	13:50
7	M3686-01DUP	1	Solid	20.03	20	21.6	12.16	09/09/2021	13:51
8	CCV2	1	Water	NA	NA	20.4	12.02	09/09/2021	14:00

LB 116215

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M3686 WorkList ID : 152107 Department : Wet-Chemistry Date : 09-09-2021 13:36:04

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/09/2021	Solid	M3686-01	pH	Cool 4 deg C	TULL02	H33	MAIN-GUARD-HOUSE-JI	09/09/2021	9045D

Date/Time 09/09/21 13:38
Raw Sample Received by: JP (WJC)
Raw Sample Relinquished by: CD SM

Date/Time 09/09/21 14:38
Raw Sample Received by: CP SM
Raw Sample Relinquished by: JP (WJC)