


LB116220

7470A

INSTRUMENT ID: CVL



Linear

A= 0.0000e+000

B= 2.5652e-004

C= -1.9876e-002

Rho= 0.9996662

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.011	-0.011	35	0.000	35					-8
0.2	0.200	0.184	-0.016	796	0.0 %	796					3
2.5	2.500	2.566	0.066	10080	0.0 %	10080					-2
5.0	5.000	4.881	-0.119	19105	0.0 %	19105					2
7.5	7.500	7.667	0.167	29967	0.0 %	29967					-1
10.0	10.000	9.912	-0.088	38719	0.0 %	38719					

LB116220 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 09 Sep 2021 14:19:47

Sample ID	Extended ID	μ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	50	35	-	0.0007470A	PPB		09 Sep 2021 14:24:42	S	Std
0.2 - 1	50.2	796	-	0.2007470A	PPB		09 Sep 2021 14:27:00	S	Std
2.5 - 1	52.5	10080	-	2.5007470A	PPB		09 Sep 2021 14:29:16	S	Std
5.0 - 1	55	19105	-	5.0007470A	PPB		09 Sep 2021 14:31:33	S	Std
7.5 - 1	57.5	29967	-	7.5007470A	PPB		09 Sep 2021 14:33:49	S	Std
10.0 - 1	57.5	38719	-	10.0007470A	PPB		09 Sep 2021 14:36:05	S	Std
ICV87 - 1	ICV87	15253	3.8929	-7470A	PPB		09 Sep 2021 14:38:50	U	SMPL
ICB87 - 1	ICB87	-25	-0.0263	-7470A	PPB		09 Sep 2021 14:41:04	U	SMPL
CCV18 - 1	CCV18	18996	4.8530	-7470A	PPB		09 Sep 2021 14:43:19	U	SMPL
CCB18 - 1	CCB18	-40	-0.0301	-7470A	PPB		09 Sep 2021 14:45:34	U	SMPL
CRA - 1	CRA	786	0.1818	-7470A	PPB		09 Sep 2021 14:47:49	U	SMPL
HighStd - 1	HighStd	38242	9.7900	-7470A	PPB		09 Sep 2021 14:50:05	U	SMPL
ChkStd - 1	ChkStd	28366	7.2566	-7470A	PPB		09 Sep 2021 14:52:20	U	SMPL
PB138997BL - 1	PBW	-59	-0.0350	-7470A	PPB		09 Sep 2021 14:54:37	U	SMPL
PB138997BS - 1	LCSW	14521	3.7051	-7470A	PPB		09 Sep 2021 14:56:53	U	SMPL
M3668-02 - 1	SP-1	-26	-0.0265	-7470A	PPB		09 Sep 2021 14:59:09	U	SMPL
M3669-01 - 1	UTILITY-POLES-BRIGE-PIERS-DECKING-RR-TIES	20	-0.0147	-7470A	PPB		09 Sep 2021 15:01:26	U	SMPL
M3670-01 - 1	SAND-BLAST-MEDIA	57	-0.0053	-7470A	PPB		09 Sep 2021 15:03:43	U	SMPL
M3671-01 - 1	FILTER-MEDIA-FILTERS	42	-0.0091	-7470A	PPB		09 Sep 2021 15:06:00	U	SMPL
M3681-03 - 1	BS-STOCK	40	-0.0096	-7470A	PPB		09 Sep 2021 15:08:15	U	SMPL
CCV19 - 1	CCV19	19210	4.9079	-7470A	PPB		09 Sep 2021 15:10:30	U	SMPL
CCB19 - 1	CCB19	-34	-0.0286	-7470A	PPB		09 Sep 2021 15:12:45	U	SMPL
M3681-03DUP - 1	BS-STOCKDUP	25	-0.0135	-7470A	PPB		09 Sep 2021 15:15:00	U	SMPL
M3681-03MS - 1	BS-STOCKMS	13565	3.4598	-7470A	PPB		09 Sep 2021 15:17:16	U	SMPL
M3681-03MSD - 1	BS-STOCKMSD	12680	3.2328	-7470A	PPB		09 Sep 2021 15:19:32	U	SMPL
PB138998BL - 1	PBW	5	-0.0186	-7470A	PPB		09 Sep 2021 15:21:48	U	SMPL
PB138998BS - 1	LCSW	14469	3.6917	-7470A	PPB		09 Sep 2021 15:26:40	U	SMPL
M3681-04 - 1	WATER-PIT	29	-0.0124	-7470A	PPB		09 Sep 2021 15:28:56	U	SMPL
M3681-04DUP - 1	WATER-PITDUP	65	-0.0032	-7470A	PPB		09 Sep 2021 15:31:12	U	SMPL
M3681-04MS - 1	WATER-PITMS	14781	3.7718	-7470A	PPB		09 Sep 2021 15:33:29	U	SMPL
M3681-04MSD - 1	WATER-PITMSD	14101	3.5973	-7470A	PPB		09 Sep 2021 15:35:46	U	SMPL
CCV20 - 1	CCV20	19398	4.9561	-7470A	PPB		09 Sep 2021 15:38:01	U	SMPL
CCB20 - 1	CCB20	-40	-0.0301	-7470A	PPB		09 Sep 2021 15:40:17	U	SMPL
PB138997TB - 1	PB138998TB	19	-0.0150	-7470A	PPB		09 Sep 2021 15:42:32	U	SMPL
M3681-03LX5 - 1		25	-0.0135	-7470A	PPB		09 Sep 2021 15:44:48	U	SMPL
M3681-03A - 1		15634	3.9906	-7470A	PPB		09 Sep 2021 15:47:04	U	SMPL
M3681-04LX5 - 1		-14	-0.0235	-7470A	PPB		09 Sep 2021 15:49:19	U	SMPL
M3681-04A - 1		14967	3.8195	-7470A	PPB		09 Sep 2021 15:51:35	U	SMPL
CCV21 - 1	CCV21	19536	4.9915	-7470A	PPB		09 Sep 2021 15:53:50	U	SMPL
CCB21 - 1	CCB21	-24	-0.0260	-7470A	PPB		09 Sep 2021 15:56:06	U	SMPL