



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB116224

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	09/09/2021	15:50
2	M3681-04		1	10	100.00	>212.0	>212.0	09/09/2021	16:10
3	M3681-04DUP		1	10	100.00	>212.0	>212.0	09/09/2021	16:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP9921 WorkList ID : 152116 Department : Wet-Chemistry Date : 09-09-2021 14:09:56

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/13/2021	Water	M3681-04	Flash Point	Cool 4 deg C	PSEG03	H33	WATER-PIT	09/08/2021	1010B

Date/Time 09/09/21 15:30
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: RJ CEXT-106b

Date/Time 09/09/21 16:40
 Raw Sample Received by: RJ CEXT-106b
 Raw Sample Relinquished by: KS (WC)