



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB116230

Slope : 98.4

BalanceID: Wc Sc-06

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	09/09/2021	16:30
2	CAL2	1	Water	NA	NA	20.3	7.00	09/09/2021	16:31
3	CAL3	1	Water	NA	NA	20.3	10.02	09/09/2021	16:32
4	ICV	1	Water	NA	NA	20.4	6.99	09/09/2021	16:35
5	CCV1	1	Water	NA	NA	20.4	2.01	09/09/2021	16:40
6	M3689-01	1	Solid	20.03	20	23.0	5.43	09/09/2021	17:05
7	M3689-01DUP	1	Solid	20.04	20	23.1	5.44	09/09/2021	17:06
8	CCV2	1	Water	NA	NA	20.4	12.03	09/09/2021	17:10

WORKLIST(Hardcopy Internal Chain)

VB 116230

WorkList Name : PH M3689 S WorkList ID : 152121 Department : Wet-Chemistry Date : 09-09-2021 16:26:14

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/09/2021	Solid	M3689-01	pH	Cool 4 deg C	PSEG03	H33	256	09/09/2021	9045D

Date/Time 09/09/21 16:25
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: JP (WC)

Date/Time 09/09/21 17:00
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: JP (WC)