



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatelANALYST: jigneshDate: 09/13/2021OVEN IN TEMP1: 104 °COVEN OUT TEMP1: 103 °CDATE & TIME IN1: 09/10/2021 16:00 DATE & TIME OUT1: 09/13/2021 08:00OVEN IN TEMP2: 103 °COVEN OUT TEMP2: 103 °CDATE & TIME IN2: 09/13/2021 08:30 DATE & TIME OUT2: 09/13/2021 10:00Run Number: LB116239BalanceID: Wc Sc-06OvenID: Wc Oven-01

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB116239BL	LB116239BL	85.4637	85.4637	100	85.4638	85.4638	85.4638	0.0001	1
2	M3686-03	MAIN-GUARD-HOUSE-JET-GROUT	90.0428	90.0428	100	90.2325	90.2325	90.2325	0.1897	1897
3	M3686-03DUP	MAIN-GUARD-HOUSE-JET-GROUTDU	90.7356	90.7356	100	90.9206	90.9206	90.9206	0.1850	1850

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result mg/L} = ((C - B) / A) * 1000 * 1000$$

VB116239

WORKLIST(Hardcopy Internal Chain)

WorkList Name : astm ts n3688 WorkList ID : 152147 Department : Wet-Chemistry Date : 09-10-2021 11:47:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/14/2021	Solid	M3686-03	ASTM TS	Cool 4 deg C	TULL02	H33	MAIN-GUARD-HOUSE-JI	09/09/2021	SM2540 B

Date/Time 09-10-21 12:30
Raw Sample Received by: JP (wlc)
Raw Sample Relinquished by: JP (wlc)

Date/Time 09-10-21 16:20
Raw Sample Received by: JP (wlc)
Raw Sample Relinquished by: JP (wlc)