



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB116246

Slope : 98.3

BalanceID: Wc Sc-06

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	09/10/2021	13:35
2	CAL2	1	Water	NA	NA	20.1	7.00	09/10/2021	13:36
3	CAL3	1	Water	NA	NA	20.2	10.03	09/10/2021	13:37
4	ICV	1	Water	NA	NA	20.3	6.98	09/10/2021	13:39
5	CCV1	1	Water	NA	NA	20.3	2.01	09/10/2021	13:40
6	M3709-02	1	Solid	20.03	20	20.3	5.56	09/10/2021	13:42
7	M3709-02DUP	1	Solid	20.04	20	5.6	5.57	09/10/2021	13:43
8	M3709-05	1	Solid	20.05	20	20.5	5.82	09/10/2021	13:45
9	CCV2	1	Water	NA	NA	20.4	12.02	09/10/2021	13:47

LB 116246

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY M3709 WorkList ID : 152157 Department : Wet-Chemistry Date : 09-10-2021 13:04:43

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/09/2021	Solid	M3709-02	☒ Corrosivity	Cool 4 deg C	NOBI03	H41	250EM-DISPOSAL-1	09/09/2021	9045D
09/09/2021	Solid	M3709-05	☒ Corrosivity	Cool 4 deg C	NOBI03	H41	250EM-DISPOSAL-2	09/09/2021	9045D

Date/Time 09-10-21 13:30
Raw Sample Received by: JP/WCJ
Raw Sample Relinquished by: CP SM

Date/Time 09-10-21 16:30
Raw Sample Received by: CP SM
Raw Sample Relinquished by: JP/WCJ