



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB116256

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.6	09/10/2021	14:20
2	M3688-01		1	10	100.00	>212.0	>212.0	09/10/2021	14:45
3	M3688-01DUP		1	10	100.00	>212.0	>212.0	09/10/2021	15:10
4	M3689-02		1	12	100.00	>212.0	>212.0	09/10/2021	15:35
5	M3689-03		1	10	100.00	>212.0	>212.0	09/10/2021	15:58
6	M3689-05		1	12	100.00	>212.0	>212.0	09/10/2021	16:25

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB116256

WorkList Name : fp91021

WorkList ID : 152152

Department : Wet-Chemistry

Date : 09-10-2021 12:31:32

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/14/2021	Water	M3688-01	Flash Point	Cool 4 deg C	PSEG03	H33	A688	09/09/2021	1010B
09/16/2021	Water	M3689-02	Flash Point	Cool 4 deg C	PSEG03	H33	253	09/09/2021	1010B
09/16/2021	Water	M3689-03	Flash Point	Cool 4 deg C	PSEG03	H33	257	09/09/2021	1010B
09/16/2021	Water	M3689-05	Flash Point	Cool 4 deg C	PSEG03	H33	258	09/09/2021	1010B

Date/Time 09/10/21 14:00
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: TP (WC)

Date/Time 09/10/21 16:30
 Raw Sample Received by: TP (WC)
 Raw Sample Relinquished by: KS (WC)