



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB116257

Analyst By : ketankumar

Supervisor Review By : jignesh

Slope : 98.4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	09/10/2021	16:20
2	CAL2	1	Water	NA	NA	20.2	7.00	09/10/2021	16:22
3	CAL3	1	Water	NA	NA	20.3	10.02	09/10/2021	16:25
4	ICV	1	Water	NA	NA	20.3	6.99	09/10/2021	16:27
5	CCV1	1	Water	NA	NA	20.3	2.01	09/10/2021	16:28
6	M3706-01	1	Water	NA	NA	21.6	4.95	09/10/2021	16:30
7	M3706-01DUP	1	Water	NA	NA	21.6	4.96	09/10/2021	16:31
8	M3706-02	1	Water	NA	NA	21.7	4.98	09/10/2021	16:33
9	M3706-03	1	Water	NA	NA	21.6	5.66	09/10/2021	16:35
10	M3706-04	1	Water	NA	NA	21.6	5.76	09/10/2021	16:37
11	M3706-05	1	Water	NA	NA	21.6	5.30	09/10/2021	16:38
12	M3706-06	1	Water	NA	NA	21.5	5.10	09/10/2021	16:40
13	M3706-07	1	Water	NA	NA	21.5	7.76	09/10/2021	16:42
14	M3706-08	1	Water	NA	NA	21.5	4.80	09/10/2021	16:43
15	M3716-05	1	Water	NA	NA	20.4	6.12	09/10/2021	16:44
16	CCV2	1	Water	NA	NA	20.4	12.02	09/10/2021	16:45

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph91021

WorkList ID : 152163

Department : Wet-Chemistry

Date : 09-10-2021 16:12:21

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/10/2021	Water	M3706-01	✓ pH	Cool 4 deg C	PSEG03	H41	33436	09/10/2021	9040C
09/10/2021	Water	M3706-02	✓ pH	Cool 4 deg C	PSEG03	H41	33437	09/10/2021	9040C
09/10/2021	Water	M3706-03	✓ pH	Cool 4 deg C	PSEG03	H41	33438	09/10/2021	9040C
09/10/2021	Water	M3706-04	✓ pH	Cool 4 deg C	PSEG03	H41	33439	09/10/2021	9040C
09/10/2021	Water	M3706-05	✓ pH	Cool 4 deg C	PSEG03	H41	33440	09/10/2021	9040C
09/10/2021	Water	M3706-06	✓ pH	Cool 4 deg C	PSEG03	H41	33441	09/10/2021	9040C
09/10/2021	Water	M3706-07	✓ pH	Cool 4 deg C	PSEG03	H41	33443	09/10/2021	9040C
09/10/2021	Water	M3706-08	✓ pH	Cool 4 deg C	PSEG03	H41	33442	09/10/2021	9040C
09/10/2021	Water	M3716-05	✓ pH	Cool 4 deg C	MAJO01	H41	DRAIN-WATER-TANK-1	09/10/2021	9040C

Date/Time

09.10.21 16:15

Raw Sample Received by:

J.P. (S.M.)

Raw Sample Relinquished by:

Date/Time

09.10.21

Raw Sample Received by:

J.P. (S.M.)

Raw Sample Relinquished by:

J.P. (S.M.)