


LB116264

7470A

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.4980e-004

C= 7.8985e-003

Rho= 0.9995408

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Std
0.0	0.000	0.025	0.025	67	0.000	67					0.00
0.2	0.200	0.209	0.009	805	0.0 %	805					4
2.5	2.500	2.432	-0.068	9705	0.0 %	9705					-3
5.0	5.000	5.150	0.150	20585	0.0 %	20585					3
7.5	7.500	7.300	-0.200	29191	0.0 %	29191					-3
10.0	10.000	10.084	0.084	40339	0.0 %	40339					1

LB116264 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 13 Sep 2021 12:11:49					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	67	-	0.0007470A	PPB		13 Sep 2021 12:49:00	S	Std
0.2 - 1	50.2	805	-	0.20007470A	PPB		13 Sep 2021 12:51:18	S	Std
2.5 - 1	52.5	9705	-	2.50007470A	PPB		13 Sep 2021 12:53:34	S	Std
5.0 - 1	55	20585	-	5.00007470A	PPB		13 Sep 2021 12:55:51	S	Std
7.5 - 1	57.5	29191	-	7.50007470A	PPB		13 Sep 2021 12:58:08	S	Std
10.0 - 1	510	40339	-	10.00007470A	PPB		13 Sep 2021 13:00:23	S	Std
ICV88 - 1	ICV88	15372	3.8478	-7470A	PPB		13 Sep 2021 13:03:14	U	SMPL
ICB88 - 1	ICB88	11	0.0106	-7470A	PPB		13 Sep 2021 13:05:29	U	SMPL
CCV21 - 1	CCV21	20403	5.1045	-7470A	PPB		13 Sep 2021 13:07:44	U	SMPL
CCB21 - 1	CCB21	9	0.0101	-7470A	PPB		13 Sep 2021 13:09:59	U	SMPL
CRA - 1	CRA	788	0.2047	-7470A	PPB		13 Sep 2021 13:12:13	U	SMPL
HighStd - 1	HighStd	40092	10.0228	-7470A	PPB		13 Sep 2021 13:14:28	U	SMPL
ChkStd - 1	ChkStd	27674	6.9208	-7470A	PPB		13 Sep 2021 13:16:44	U	SMPL
PB139054BL - 1	PBW	13	0.0111	-7470A	PPB		13 Sep 2021 13:18:59	U	SMPL
PB139054BS - 1	LCSW	15170	3.7973	-7470A	PPB		13 Sep 2021 13:21:16	U	SMPL
M3686-02 - 1	MAIN-GUARD-HOUSE-JET-GROUT	8	0.0099	-7470A	PPB		13 Sep 2021 13:23:32	U	SMPL
M3686-02DUP - 1	MAIN-GUARD-HOUSE-JET-GROUTDUP	87	0.0296	-7470A	PPB		13 Sep 2021 13:25:48	U	SMPL
M3686-02MS - 1	MAIN-GUARD-HOUSE-JET-GROUTMS	14660	3.6699	-7470A	PPB		13 Sep 2021 13:28:05	U	SMPL
M3686-02MSD - 1	MAIN-GUARD-HOUSE-JET-GROUTMSD	15557	3.8940	-7470A	PPB		13 Sep 2021 13:30:22	U	SMPL
M3689-01 - 1	256	204	0.0589	-7470A	PPB		13 Sep 2021 13:32:37	U	SMPL
CCV22 - 1	CCV22	20014	5.0073	-7470A	PPB		13 Sep 2021 13:34:52	U	SMPL
CCB22 - 1	CCB22	0	0.0079	-7470A	PPB		13 Sep 2021 13:37:07	U	SMPL
PB139055BL - 1	PBW	37	0.0171	-7470A	PPB		13 Sep 2021 13:42:32	U	SMPL
PB139055BS - 1	LCSW	15670	3.9222	-7470A	PPB		13 Sep 2021 13:44:46	U	SMPL
M3689-02 - 1	253	1	0.0081	-7470A	PPB		13 Sep 2021 13:47:02	U	SMPL
M3689-02DUP - 1	253DUP	82	0.0284	-7470A	PPB		13 Sep 2021 13:49:17	U	SMPL
M3689-02MS - 1	253MS	14590	3.6524	-7470A	PPB		13 Sep 2021 13:51:33	U	SMPL
M3689-02MSD - 1	253MSD	14815	3.7086	-7470A	PPB		13 Sep 2021 13:53:49	U	SMPL
M3689-03 - 1	257	259	0.0726	-7470A	PPB		13 Sep 2021 13:56:05	U	SMPL
M3689-05 - 1	258	64	0.0239	-7470A	PPB		13 Sep 2021 13:58:22	U	SMPL
PB139056BL - 1	PBW	147	0.0446	-7470A	PPB		13 Sep 2021 14:00:39	U	SMPL
PB139056BS - 1	LCSW	13731	3.4379	-7470A	PPB		13 Sep 2021 14:02:55	U	SMPL
CCV23 - 1	CCV23	19969	4.9961	-7470A	PPB		13 Sep 2021 14:05:10	U	SMPL
CCB23 - 1	CCB23	-1	0.0076	-7470A	PPB		13 Sep 2021 14:07:26	U	SMPL
M3709-02 - 1	250EM-DISPOSAL-1	85	0.0291	-7470A	PPB		13 Sep 2021 14:09:41	U	SMPL
M3709-02DUP - 1	250EM-DISPOSAL-1DUP	97	0.0321	-7470A	PPB		13 Sep 2021 14:11:56	U	SMPL
M3709-02MS - 1	250EM-DISPOSAL-1MS	14212	3.5580	-7470A	PPB		13 Sep 2021 14:14:11	U	SMPL
M3709-02MSD - 1	250EM-DISPOSAL-1MSD	14387	3.6017	-7470A	PPB		13 Sep 2021 14:16:27	U	SMPL
M3709-05 - 1	250EM-DISPOSAL-2	16	0.0119	-7470A	PPB		13 Sep 2021 14:18:43	U	SMPL
PB139057BL - 1	PBW	46	0.0194	-7470A	PPB		13 Sep 2021 14:21:00	U	SMPL
PB139057BS - 1	LCSW	15017	3.7591	-7470A	PPB		13 Sep 2021 14:23:16	U	SMPL
M3684-22 - 1	FB-210908	14	0.0114	-7470A	PPB		13 Sep 2021 14:25:33	U	SMPL
M3684-24 - 1	FB-210909	66	0.0244	-7470A	PPB		13 Sep 2021 14:27:50	U	SMPL
M3688-01 - 1	A688	189	0.0551	-7470A	PPB		13 Sep 2021 14:30:06	U	SMPL
CCV24 - 1	CCV24	19925	4.9851	-7470A	PPB		13 Sep 2021 14:32:22	U	SMPL
CCB24 - 1	CCB24	-2	0.0074	-7470A	PPB		13 Sep 2021 14:34:37	U	SMPL
M3688-01DUP - 1	A688DUP	191	0.0556	-7470A	PPB		13 Sep 2021 14:36:53	U	SMPL
M3688-01MS - 1	A688MS	14111	3.5328	-7470A	PPB		13 Sep 2021 14:39:08	U	SMPL
M3688-01MSD - 1	A688MSD	14572	3.6479	-7470A	PPB		13 Sep 2021 14:41:24	U	SMPL
PB139054TB - 1	PB139004TB	9	0.0101	-7470A	PPB		13 Sep 2021 14:43:39	U	SMPL
PB139055TB - 1	PB139011TB	107	0.0346	-7470A	PPB		13 Sep 2021 14:45:55	U	SMPL
PB139056TB - 1	PB139033TB	72	0.0259	-7470A	PPB		13 Sep 2021 14:48:11	U	SMPL
CCV25 - 1	CCV25	19751	4.9416	-7470A	PPB		13 Sep 2021 14:50:27	U	SMPL
CCB25 - 1	CCB25	26	0.0144	-7470A	PPB		13 Sep 2021 14:52:45	U	SMPL
M3686-02LX5 - 1		76	0.0269	-7470A	PPB		13 Sep 2021 14:57:36	U	SMPL
M3686-02A - 1		15200	3.8048	-7470A	PPB		13 Sep 2021 14:59:51	U	SMPL
M3689-02LX5 - 1		-13	0.0047	-7470A	PPB		13 Sep 2021 15:02:07	U	SMPL
M3689-02A - 1		15033	3.7631	-7470A	PPB		13 Sep 2021 15:04:22	U	SMPL
M3709-02LX5 - 1		5	0.0091	-7470A	PPB		13 Sep 2021 15:06:38	U	SMPL
M3709-02A - 1		15349	3.8420	-7470A	PPB		13 Sep 2021 15:08:53	U	SMPL
M3688-01LX5 - 1		-5	0.0066	-7470A	PPB		13 Sep 2021 15:11:09	U	SMPL
M3688-01A - 1		15234	3.8133	-7470A	PPB		13 Sep 2021 15:13:24	U	SMPL
CCV26 - 1	CCV26	20018	5.0083	-7470A	PPB		13 Sep 2021 15:15:40	U	SMPL
CCB26 - 1	CCB26	-32	-0.0001	-7470A	PPB		13 Sep 2021 15:17:57	U	SMPL